

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122319\
 Data File : PL055186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2019 16:41
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 00:49:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	4.049	550.5E6	141.3E6	52.349	53.488
28)	SA Decachlor...	8.314	9.147	440.0E6	108.7E6	47.416	49.263
Target Compounds							
2)	A alpha-BHC	3.978	4.440	910.3E6	201.9E6	52.326	54.831
3)	MA gamma-BHC...	4.264	4.726	828.9E6	184.5E6	50.272	53.978
4)	MA Heptachlor	4.569	5.235	837.7E6	167.3E6	51.309	53.186
5)	MB Aldrin	4.820	5.540	782.9E6	153.5E6	51.270	53.307
6)	B beta-BHC	4.515	4.896	348.1E6	82377753	50.758	50.163
7)	B delta-BHC	4.723	5.114	791.2E6	162.2E6	52.270	57.988
8)	B Heptachlo...	5.265	5.927	718.9E6	139.3E6	50.312	54.637
9)	A Endosulfan I	5.607	6.285	663.3E6	129.2E6	50.199	52.163
10)	B gamma-Chl...	5.493	6.162	733.3E6	139.8E6	50.331	52.678
11)	B alpha-Chl...	5.552	6.235	702.0E6	137.4E6	49.858	52.395
12)	B 4,4'-DDE	5.715	6.391	668.6E6	128.8E6	50.169	52.817
13)	MA Dieldrin	5.851	6.543	702.9E6	134.2E6	50.349	52.573
14)	MA Endrin	6.110	6.762	621.0E6	110.8E6	48.862	51.630
15)	B Endosulfa...	6.382	6.972	560.2E6	118.3E6	47.730	49.902
16)	A 4,4'-DDD	6.233	6.884	494.1E6	102.3E6	51.440	53.830
17)	MA 4,4'-DDT	6.474	7.183	503.9E6	100.4E6	50.051	51.162
18)	B Endrin al...	6.551	7.096	467.9E6	99174780	48.044	51.388
19)	B Endosulfa...	6.764	7.321	498.1E6	106.6E6	49.520	51.055
20)	A Methoxychlor	7.019	7.643	241.9E6	54218492	50.190	50.236
21)	B Endrin ke...	7.259	7.794	537.9E6	113.7E6	51.313	51.048
22)	Mirex	7.451	8.251	367.2E6	82620225	48.214	49.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122319\
Data File : PL055186.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2019 16:41
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 00:49:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

