

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122320\
 Data File : PL063778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2020 13:40
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 13:55:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 14:02:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.437	4.110	170.0E6	143.7E6	56.711	56.279
28)	SA Decachlor...	8.174	9.265	120.0E6	87442776	43.553	44.351
Target Compounds							
2)	A alpha-BHC	3.871	4.512	287.8E6	209.3E6	57.658	58.087
3)	MA gamma-BHC...	4.152	4.803	267.6E6	196.0E6	58.533	57.443
4)	MA Heptachlor	4.452	5.315	241.5E6	172.9E6	57.345	54.150
5)	MB Aldrin	4.698	5.625	232.3E6	156.5E6	58.159	54.682
6)	B beta-BHC	4.400	4.978	107.4E6	90546455	55.582	51.584
7)	B delta-BHC	4.603	5.200	265.2E6	187.2E6	58.505	56.952
8)	B Heptachlo...	5.139	6.018	217.2E6	148.7E6	57.801	53.577
9)	A Endosulfan I	5.475	6.381	195.9E6	131.6E6	56.585	52.326
10)	B gamma-Chl...	5.365	6.253	218.6E6	145.7E6	57.494	52.676
11)	B alpha-Chl...	5.423	6.327	213.6E6	144.9E6	56.932	52.503
12)	B 4,4'-DDE	5.590	6.483	205.6E6	127.1E6	59.813	51.468
13)	MA Dieldrin	5.717	6.641	204.0E6	134.5E6	56.293	51.633
14)	MA Endrin	5.972	6.864	172.0E6	91594216	59.012	47.199
15)	B Endosulfa...	6.244	7.075	175.7E6	115.9E6	53.299	49.145
16)	A 4,4'-DDD	6.103	6.980	163.4E6	105.8E6	55.853	50.033
17)	MA 4,4'-DDT	6.343	7.280	142.6E6	90610195	53.825	46.575
18)	B Endrin al...	6.412	7.200	143.5E6	96276368	51.868	47.756
19)	B Endosulfa...	6.625	7.424	166.3E6	108.0E6	52.810	48.035
20)	A Methoxychlor	6.890	7.738	67604348	48301045	51.458	45.781
21)	B Endrin ke...	7.111	7.903	174.1E6	116.8E6	50.091	47.365

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122320\
Data File : PL063778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2020 13:40
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 13:55:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
Quant Title : GC Extractables
QLast Update : Tue Dec 22 14:02:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

