

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX122322.M
 Title : GC Extractables
 Last Update : Sun Dec 25 19:53:24 2022
 Response Via : Initial Calibration

Calibration Files

500 =PL079885.D 1000 =PL079883.D 250 =PL079886.D
 750 =PL079884.D 100 =PL079887.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.377	2.376	2.382	2.286	2.115	2.307	E6	4.98
2)	Toxaphene-1	1.818	1.740	1.780	1.698	1.400	1.687	E4	9.88
3)	Toxaphene-2	6.912	6.656	7.074	6.413	6.202	6.651	E4	5.35
4)	Toxaphene-3	5.483	5.307	5.588	5.189	4.699	5.253	E4	6.58
5)	Toxaphene-4	3.894	3.710	3.907	3.608	3.284	3.680	E4	6.92
6)	Toxaphene-5	3.526	3.501	3.653	3.383	3.220	3.457	E4	4.73
7) SA	Decachlorobiphenyl	2.020	1.920	2.080	1.884	1.981	1.977	E6	3.95

Signal #2 Calibration Files

500 =PL079885.D 1000 =PL079883.D 250 =PL079886.D
 750 =PL079884.D 100 =PL079887.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.896	2.814	2.913	2.711	2.547	2.777	E6	5.44
2)	Toxaphene-1	1.639	1.556	1.695	1.512	1.492	1.579	E4	5.46
3)	Toxaphene-2	2.515	2.371	2.683	2.311	2.196	2.415	E4	7.81
4)	Toxaphene-3	7.346	7.124	7.357	6.783	5.769	6.876	E4	9.61
5)	Toxaphene-4	1.023	0.992	1.039	0.959	0.907	0.984	E5	5.40
6)	Toxaphene-5	6.559	6.416	6.654	6.105	5.293	6.205	E4	8.87
7) SA	Decachlorobiphenyl	2.243	2.121	2.333	2.070	2.281	2.210	E6	4.99

(#) = Out of Range