

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122523\
 Data File : PL087426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Dec 2023 15:33
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 12/26/2023
 Supervised By :Ankita Jodhani 12/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 15:48:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121523.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 16 04:03:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.421	2.677	189.2E6	79869970	50.261	47.449
2)	SA Decachlor...	8.979	7.863	175.1E6	81379080	51.347	48.183m
Target Compounds							
2)	A alpha-BHC	3.877	3.179	278.9E6	117.4E6	51.514	47.787m
3)	MA gamma-BHC...	4.211	3.512	266.3E6	113.3E6	50.662	46.759
4)	MA Heptachlor	4.802	3.854	276.8E6	112.8E6	51.975m	47.201
5)	MB Aldrin	5.146	4.134	255.4E6	109.2E6	49.057	45.538
6)	B beta-BHC	4.410	3.812	111.1E6	46465271	48.370	44.636
7)	B delta-BHC	4.655	4.042	266.6E6	112.1E6	50.988m	45.817
8)	B Heptachlo...	5.577	4.641	228.1E6	99187690	48.392	44.573
9)	A Endosulfan I	5.963	5.012	212.2E6	94748888	47.521	43.841
10)	B gamma-Chl...	5.835	4.895	231.9E6	100.4E6	47.827	43.879
11)	B alpha-Chl...	5.915	4.959	229.4E6	101.4E6	47.740	43.945
12)	B 4,4'-DDE	6.096	5.157	198.4E6	86142242	47.869	43.385
13)	MA Dieldrin	6.241	5.279	219.7E6	98457008	47.342	44.196
14)	MA Endrin	6.471	5.555	189.8E6	88987695	49.319	44.591
15)	B Endosulfa...	6.693	5.853	188.9E6	85606134	47.100	43.968
16)	A 4,4'-DDD	6.616	5.715	157.9E6	72940472	47.776	44.053
17)	MA 4,4'-DDT	6.934	5.968	181.9E6	81866572	48.381	44.388
18)	B Endrin al...	6.824	6.035	141.1E6	65165668	48.671	44.402
19)	B Endosulfa...	7.061	6.260	184.7E6	82892852	47.042	44.003
20)	A Methoxychlor	7.419	6.554	95972413	45292705	49.219	44.344
21)	B Endrin ke...	7.545	6.764	193.0E6	95634008	47.434m	44.926m
22)	Mirex	8.024	6.947	157.2E6	79820001	46.688	44.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122523\
Data File : PL087426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Dec 2023 15:33
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/26/2023
Supervised By :Ankita Jodhani 12/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 25 15:48:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL121523.M
Quant Title : GC Extractables
QLast Update : Sat Dec 16 04:03:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

