

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122719\
 Data File : PL055312.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2019 12:04
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:29:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.534	4.048	515.0E6	136.8E6	48.974	51.770
2)	SA Decachlor...	8.311	9.145	422.4E6	106.5E6	45.520	48.238
Target Compounds							
2)	A alpha-BHC	3.977	4.439	853.2E6	196.6E6	49.041	53.392
3)	MA gamma-BHC...	4.263	4.724	781.1E6	180.5E6	47.372	52.814
4)	MA Heptachlor	4.568	5.233	771.6E6	159.6E6	47.257	50.737
5)	MB Aldrin	4.819	5.538	751.6E6	152.8E6	49.221	53.055
6)	B beta-BHC	4.514	4.895	331.1E6	81186357	48.279	49.438
7)	B delta-BHC	4.721	5.113	753.5E6	164.6E6	49.783	58.877
8)	B Heptachlo...	5.265	5.925	694.6E6	140.2E6	48.611	54.998
9)	A Endosulfan I	5.606	6.283	644.3E6	128.7E6	48.762	51.970
10)	B gamma-Chl...	5.492	6.160	707.7E6	139.1E6	48.573	52.390
11)	B alpha-Chl...	5.551	6.233	680.2E6	136.2E6	48.314	51.945
12)	B 4,4'-DDE	5.714	6.390	651.3E6	127.8E6	48.872	52.418
13)	MA Dieldrin	5.850	6.541	686.3E6	133.0E6	49.159	52.096
14)	MA Endrin	6.109	6.759	566.1E6	103.5E6	44.538	48.221
15)	B Endosulfa...	6.381	6.970	561.3E6	118.0E6	47.823	49.767
16)	A 4,4'-DDD	6.231	6.882	511.5E6	106.3E6	53.255	55.893
17)	MA 4,4'-DDT	6.473	7.181	427.9E6	87673684	42.507	44.672
18)	B Endrin al...	6.550	7.094	461.8E6	98631027	47.413	51.106
19)	B Endosulfa...	6.763	7.318	489.0E6	104.6E6	48.614	50.076
20)	A Methoxychlor	7.017	7.641	201.4E6	48096452	41.790	44.564
21)	B Endrin ke...	7.257	7.792	547.4E6	117.8E6	52.220	52.913
22)	Mirex	7.448	8.248	355.7E6	82805770	46.697	49.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122719\
Data File : PL055312.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2019 12:04
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 00:29:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

