

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122818\
 Data File : PL043491.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Dec 2018 12:16
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 1/2/2019 8:47:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 22:28:00 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 07 00:20:08 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.464	4.100	329.1E6	166.6E6	53.657	56.194
2)	SA Decachlor...	8.200	9.230	296.2E6	124.7E6	45.355	47.815
Target Compounds							
2)	A alpha-BHC	3.904	4.495	508.9E6	245.3E6	53.830	54.220
3)	MA gamma-BHC...	4.187	4.784	461.6E6	226.5E6	53.457	54.012
4)	MA Heptachlor	4.482	5.299	454.9E6	216.2E6	51.265	52.786
5)	MB Aldrin	4.728	5.608	442.9E6	219.2E6	52.242	53.701
6)	B beta-BHC	4.441	4.953	189.0E6	96294523	51.234	56.880
7)	B delta-BHC	4.643	5.174	433.0E6	216.2E6	53.468	50.313
8)	B Heptachlo...	5.170	5.995	400.0E6	199.0E6	50.459	52.488m
9)	A Endosulfan I	5.508	6.356	367.3E6	176.9E6	50.133	52.112
10)	B gamma-Chl...	5.396	6.231	407.6E6	195.1E6	50.560	53.105
11)	B alpha-Chl...	5.454	6.304	396.4E6	189.8E6	49.925	52.771
12)	B 4,4'-DDE	5.618	6.459	362.6E6	190.5E6	52.380m	52.557
13)	MA Dieldrin	5.750	6.615	393.0E6	189.6E6	50.542	52.293
14)	MA Endrin	6.007	6.834	330.1E6	138.1E6	45.844	44.228
15)	B Endosulfa...	6.282	7.043	319.2E6	154.4E6	47.980	50.978
16)	A 4,4'-DDD	6.136	6.951	301.8E6	152.2E6	53.209	51.912
17)	MA 4,4'-DDT	6.374	7.251	262.0E6	133.0E6	44.048	45.280
18)	B Endrin al...	6.450	7.167	258.5E6	126.0E6	46.918	50.252
19)	B Endosulfa...	6.663	7.390	320.3E6	153.5E6	47.813	51.050
20)	A Methoxychlor	6.920	7.708	145.7E6	66458442	43.418	44.061
21)	B Endrin ke...	7.150	7.865	366.4E6	155.2E6	50.671m	50.789

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122818\
Data File : PL043491.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Dec 2018 12:16
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
1/2/2019 8:47:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 22:28:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL120618.M
Quant Title : GC Extractables
QLast Update : Fri Dec 07 00:20:08 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

