

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122823\
 Data File : PL087535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Dec 2023 18:41
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 00:57:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 29 00:57:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.420	2.675	297.5E6	124.4E6	100.810	101.850
28)	SA Decachlor...	8.983	7.865	259.5E6	117.8E6	97.072	97.490
Target Compounds							
2)	A alpha-BHC	3.877	3.179	482.9E6	201.3E6	103.426	103.440
3)	MA gamma-BHC...	4.211	3.510	460.8E6	193.7E6	102.715	102.625
4)	MA Heptachlor	4.804	3.852	447.9E6	186.6E6	101.555	102.001
5)	MB Aldrin	5.147	4.133	445.2E6	187.1E6	101.927	102.469
6)	B beta-BHC	4.410	3.811	179.8E6	75218442	99.202	99.785
7)	B delta-BHC	4.657	4.041	441.9E6	188.4E6	102.763	103.161
8)	B Heptachlo...	5.577	4.640	387.9E6	164.2E6	101.190	101.231
9)	A Endosulfan I	5.965	5.012	353.7E6	151.8E6	100.618	100.622
10)	B gamma-Chl...	5.836	4.894	389.5E6	166.2E6	101.419	101.889
11)	B alpha-Chl...	5.917	4.958	384.6E6	166.6E6	100.882	101.188
12)	B 4,4'-DDE	6.098	5.157	334.9E6	145.8E6	102.075	102.695
13)	MA Dieldrin	6.243	5.279	384.6E6	168.4E6	101.741	102.226
14)	MA Endrin	6.473	5.554	321.8E6	148.9E6	101.506	101.870
15)	B Endosulfa...	6.695	5.852	312.6E6	140.4E6	100.129	101.002
16)	A 4,4'-DDD	6.618	5.715	266.8E6	120.3E6	101.603	102.065
17)	MA 4,4'-DDT	6.935	5.967	294.1E6	132.8E6	101.153	101.949
18)	B Endrin al...	6.825	6.034	245.7E6	112.4E6	99.417	99.436
19)	B Endosulfa...	7.063	6.260	299.5E6	132.4E6	99.823	100.387
20)	A Methoxychlor	7.421	6.554	155.2E6	71529532	98.334	98.379
21)	B Endrin ke...	7.549	6.766	324.1E6	156.0E6	100.061	100.088
22)	Mirex	8.027	6.947	255.7E6	126.0E6	97.294	98.292

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL122823\
Data File : PL087535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Dec 2023 18:41
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 00:57:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
Quant Title : GC Extractables
QLast Update : Fri Dec 29 00:57:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

