

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL123019\
 Data File : PL055363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Dec 2019 10:35
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 12/31/2019 11:58:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 00:50:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 21 05:21:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	4.049	500.9E6	136.1E6	47.630	51.520
28)	SA Decachlor...	8.320	9.154	435.7E6	103.8E6	46.955	47.003
Target Compounds							
2)	A alpha-BHC	3.984	4.441	829.8E6	194.9E6	47.697	52.922
3)	MA gamma-BHC...	4.271	4.727	764.9E6	180.0E6	46.390	52.656
4)	MA Heptachlor	4.577	5.236	764.5E6	159.6E6	46.824	50.724
5)	MB Aldrin	4.828	5.542	733.5E6	150.6E6	48.035	52.292
6)	B beta-BHC	4.522	4.896	313.5E6	89136497	45.723	54.279m
7)	B delta-BHC	4.730	5.115	729.6E6	161.5E6	48.204	57.748
8)	B Heptachlo...	5.273	5.930	680.0E6	136.6E6	47.587	53.575
9)	A Endosulfan I	5.615	6.288	623.9E6	126.8E6	47.217	51.193
10)	B gamma-Chl...	5.502	6.164	693.3E6	137.4E6	47.583	51.756
11)	B alpha-Chl...	5.560	6.237	669.4E6	134.5E6	47.548	51.284
12)	B 4,4'-DDE	5.723	6.394	636.4E6	126.9E6	47.753	52.066
13)	MA Dieldrin	5.859	6.545	669.3E6	132.2E6	47.944	51.788
14)	MA Endrin	6.118	6.764	550.5E6	101.2E6	43.311	47.161
15)	B Endosulfa...	6.391	6.975	544.8E6	116.3E6	46.420	49.073
16)	A 4,4'-DDD	6.241	6.887	499.0E6	105.0E6	51.955	55.254
17)	MA 4,4'-DDT	6.482	7.187	437.9E6	87549914	43.496	44.609
18)	B Endrin al...	6.559	7.100	458.0E6	96925410	47.022	50.222
19)	B Endosulfa...	6.772	7.324	490.2E6	103.8E6	48.738	49.689
20)	A Methoxychlor	7.027	7.647	214.3E6	47133813	44.458	43.672
21)	B Endrin ke...	7.266	7.798	549.9E6	115.2E6	52.465	51.722
22)	Mirex	7.458	8.256	351.4E6	79657002	46.130	48.042

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL123019\
Data File : PL055363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Dec 2019 10:35
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
12/31/2019 11:58:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 00:50:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122119.M
Quant Title : GC Extractables
QLast Update : Sat Dec 21 05:21:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

