

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL123120\
 Data File : PL063906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2020 20:17
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleID :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 1/4/2021 2:26:11 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 01 00:39:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 14:02:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.437	4.110	172.3E6	141.9E6	57.486	55.590
28)	SA Decachlor...	8.174	9.264	139.5E6	90110734	50.639	45.704
Target Compounds							
2)	A alpha-BHC	3.871	4.511	292.3E6	205.9E6	58.563	57.153
3)	MA gamma-BHC...	4.152	4.803	267.3E6	190.5E6	58.483	55.845
4)	MA Heptachlor	4.452	5.315	237.8E6	168.0E6	56.469	52.611
5)	MB Aldrin	4.697	5.625	229.3E6	154.5E6	57.403	53.978
6)	B beta-BHC	4.400	4.978	104.9E6	85969162	54.270	48.976
7)	B delta-BHC	4.603	5.199	259.2E6	177.4E6	57.184	53.952
8)	B Heptachlor...	5.138	6.018	211.2E6	144.1E6	56.229	51.910
9)	A Endosulfan I	5.474	6.380	193.3E6	129.9E6	55.841	51.633
10)	B gamma-Chl...	5.365	6.253	214.2E6	143.9E6	56.348	52.020
11)	B alpha-Chl...	5.423	6.326	207.8E6	143.3E6	55.387	51.909
12)	B 4,4'-DDE	5.588	6.482	195.2E6	128.0E6	56.774m	51.815
13)	MA Dieldrin	5.717	6.640	203.8E6	134.8E6	56.218	51.752
14)	MA Endrin	5.972	6.863	170.4E6	93909361	58.476	48.392
15)	B Endosulfa...	6.244	7.075	183.9E6	116.8E6	55.797	49.558
16)	A 4,4'-DDD	6.103	6.980	182.3E6	115.0E6	62.340	54.384
17)	MA 4,4'-DDT	6.341	7.279	134.6E6	84421705	50.776m	43.394
18)	B Endrin al...	6.412	7.200	147.2E6	97146242	53.172	48.187
19)	B Endosulfa...	6.625	7.423	175.5E6	109.3E6	55.735	48.583
20)	A Methoxychlor	6.890	7.738	67034127	45644692	51.024	43.263
21)	B Endrin ke...	7.111	7.902	194.6E6	122.3E6	55.976	49.577
22)	Mirex	7.298	8.355	137.7E6	84070078	54.850	46.606

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL123120\
 Data File : PL063906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2020 20:17
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 1/4/2021 2:26:11 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 01 00:39:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 14:02:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

