

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021819\
 Data File : PL044898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2019 16:50
 Operator : AJ\SJ
 Sample : K1514-01MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-BPM-004-CLEARBROOK-PILE-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 02:16:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.365	3.983	146.8E6	50087102	20.634	21.321
28)	SA Decachlor...	8.049	9.043	153.7E6	45016967	21.073	20.325
Target Compounds							
2)	A alpha-BHC	3.795	4.373	551.3E6	154.2E6	51.622	50.616
3)	MA gamma-BHC...	4.073	4.658	507.9E6	144.5E6	51.623	49.862
4)	MA Heptachlor	4.361	5.165	515.1E6	146.9E6	49.999	48.242
5)	MB Aldrin	4.601	5.470	477.3E6	131.1E6	49.589	47.306
6)	B beta-BHC	4.325	4.829	206.7E6	64542844	48.923	48.711
7)	B delta-BHC	4.522	5.046	403.2E6	84914513	50.092	46.438
8)	B Heptachlo...	5.039	5.856	442.2E6	109.4E6	47.528	43.747
9)	A Endosulfan I	5.371	6.212	410.2E6	115.3E6	47.003	45.558
10)	B gamma-Chl...	5.262	6.090	452.7E6	123.4E6	46.906	44.463
11)	B alpha-Chl...	5.319	6.162	444.2E6	125.0E6	46.459	45.100
12)	B 4,4'-DDE	5.486	6.321	402.4E6	108.8E6	46.400	45.576
13)	MA Dieldrin	5.611	6.470	443.2E6	119.3E6	47.223	46.252
14)	MA Endrin	5.865	6.687	403.0E6	100.2E6	46.007	44.596
15)	B Endosulfa...	6.139	6.897	381.8E6	106.8E6	45.158	46.356
16)	A 4,4'-DDD	5.998	6.810	328.5E6	89382229	45.406	46.549
17)	MA 4,4'-DDT	6.233	7.109	308.8E6	85200733	43.059	43.430
18)	B Endrin al...	6.306	7.021	305.9E6	87290566	43.640	43.024
19)	B Endosulfa...	6.518	7.243	331.3E6	93455209	45.580	44.292
20)	A Methoxychlor	6.778	7.567	168.0E6	50543792	41.870	42.306
21)	B Endrin ke...	7.001	7.713	402.7E6	106.8E6	48.396	47.036

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021819\
Data File : PL044898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2019 16:50
Operator : AJ\SJ
Sample : K1514-01MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
FK-BPM-004-CLEARBROOK-PILE-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 02:16:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

