

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031023\
 Data File : PL081412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2023 17:03
 Operator : AR\AJ
 Sample : 01853-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 20230163-COMPOSITE-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 21:27:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.332	2.662	39921245	36008789	18.477	17.792
28) SA Decachlor...	8.757	7.785	21059577	25041601	13.814	13.073

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031023\
Data File : PL081412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2023 17:03
Operator : AR\AJ
Sample : 01853-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
20230163-COMPOSITE-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 21:27:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
Quant Title : GC Extractables
QLast Update : Fri Mar 03 01:48:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.332 min
Delta R.T.: -0.003 min
Response: 39921245
Conc: 18.48 ng/ml

Instrument :
ECD_L
ClientSampleId :
20230163-COMPOSITE-2




#1 Tetrachloro-m-xylene

R.T.: 2.662 min
Delta R.T.: -0.005 min
Response: 36008789
Conc: 17.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.757 min
Delta R.T.: -0.007 min
Response: 21059577
Conc: 13.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.785 min
Delta R.T.: -0.009 min
Response: 25041601
Conc: 13.07 ng/ml