

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041921\
 Data File : PL066288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2021 13:29
 Operator : AR\AJ
 Sample : M2075-01MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-1-SMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 01:58:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 14 15:37:20 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.444	4.154	69685781	118.7E6	27.378	29.506
2)	SA Decachlor...	8.168	9.353	85625490	112.5E6	35.898	34.436
Target Compounds							
2)	A alpha-BHC	3.879	4.559	205.9E6	310.6E6	50.045	53.492
3)	MA gamma-BHC...	4.159	4.852	196.5E6	285.1E6	50.435	53.552
4)	MA Heptachlor	4.458	5.368	206.3E6	294.1E6	49.732	52.837
5)	MB Aldrin	4.703	5.680	196.2E6	257.3E6	50.106	51.578
6)	B beta-BHC	4.407	5.028	88194787	134.6E6	50.045	54.143
7)	B delta-BHC	4.608	5.251	213.1E6	278.6E6	51.660	54.180
8)	B Heptachlo...	5.141	6.075	189.0E6	230.3E6	51.016	49.865
9)	A Endosulfan I	5.475	6.439	192.8E6	218.0E6	53.845	52.095
10)	B gamma-Chl...	5.366	6.311	201.7E6	240.0E6	54.471	52.592
11)	B alpha-Chl...	5.424	6.385	198.0E6	236.3E6	53.454	52.288
12)	B 4,4'-DDE	5.589	6.540	206.4E6	222.4E6	55.320	53.737
13)	MA Dieldrin	5.716	6.700	216.9E6	227.4E6	53.543	54.418
14)	MA Endrin	5.970	6.923	187.6E6	184.6E6	53.806	57.121
15)	B Endosulfa...	6.241	7.137	178.2E6	195.5E6	54.287	56.068
16)	A 4,4'-DDD	6.101	7.040	157.2E6	181.0E6	47.746	52.898
17)	MA 4,4'-DDT	6.340	7.341	172.6E6	176.0E6	54.160	51.881
18)	B Endrin al...	6.408	7.263	146.5E6	168.0E6	56.822	60.066
19)	B Endosulfa...	6.620	7.487	183.3E6	193.7E6	53.314	54.834
20)	A Methoxychlor	6.886	7.801	91200242	94461501	50.373	52.716
21)	B Endrin ke...	7.106	7.968	190.9E6	191.8E6	51.925	52.133
22)	Mirex	7.295	8.423	147.0E6	146.8E6	51.389	51.846

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL041921\
Data File : PL066288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2021 13:29
Operator : AR\AJ
Sample : M2075-01MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-1-SMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 01:58:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL041421.M
Quant Title : GC Extractables
QLast Update : Wed Apr 14 15:37:20 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

