

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053022\
 Data File : PL075262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2022 17:07
 Operator : AR\AJ
 Sample : N2985-10
 Misc :
 ALS Vial : 82 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 P021-SS005-0006-01

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/31/2022
 Supervised By : Ankita Jodhani 05/31/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 03:23:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
 Quant Title : GC Extractables
 QLast Update : Tue May 31 02:52:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.651	4.660f	4261450	14139915	7.251	8.926m
2) SA Decachlor...	11.642	10.393	6113480	14175297	11.402	10.330

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053022\
Data File : PL075262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2022 17:07
Operator : AR\AJ
Sample : N2985-10
Misc :
ALS Vial : 82 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
P021-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/31/2022
Supervised By : Ankita Jodhani 05/31/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 03:23:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
Quant Title : GC Extractables
QLast Update : Tue May 31 02:52:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.651 min
Delta R.T.: 0.010 min
Response: 4261450
Conc: 7.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
P021-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/31/2022
Supervised By :Ankita Jodhani 05/31/2022




#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.018 min
Response: 14139915
Conc: 8.93 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.642 min
Delta R.T.: 0.000 min
Response: 6113480
Conc: 11.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.393 min
Delta R.T.: 0.000 min
Response: 14175297
Conc: 10.33 ng/ml