

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060322\
 Data File : PL075484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 22:27
 Operator : AR\AJ
 Sample : N2989-05
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 P015-SS007-0006-01

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/06/2022
 Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 00:46:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 01 10:19:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.643	4.642	8499473	22473050	14.461	14.186
2) SA Decachlor...	11.638	10.386	3925228	8282501	7.321	6.036
Target Compounds						
12) B 4,4'-DDE	8.659	7.627	321918	1317483	0.553m	0.809 #
13) MA Dieldrin	8.810	7.763	1589227	806838	2.508	0.465 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060322\
Data File : PL075484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 22:27
Operator : AR\AJ
Sample : N2989-05
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
P015-SS007-0006-01


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/06/2022
Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 00:46:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
Quant Title : GC Extractables
QLast Update : Wed Jun 01 10:19:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.643 min
Delta R.T.: 0.002 min
Response: 8499473
Conc: 14.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
P015-SS007-0006-01


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022




#1 Tetrachloro-m-xylene

R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 22473050
Conc: 14.19 ng/ml




#12 4,4'-DDE

R.T.: 8.659 min
Delta R.T.: -0.001 min
Response: 321918
Conc: 0.55 ng/ml m



#12 4,4'-DDE

R.T.: 7.627 min
Delta R.T.: -0.004 min
Response: 1317483
Conc: 0.81 ng/ml




#13 Dieldrin

R.T.: 8.810 min
Delta R.T.: -0.011 min
Response: 1589227
Conc: 2.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
P015-SS007-0006-01


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022




#13 Dieldrin

R.T.: 7.763 min
Delta R.T.: 0.003 min
Response: 806838
Conc: 0.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.638 min
Delta R.T.: -0.003 min
Response: 3925228
Conc: 7.32 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: -0.005 min
Response: 8282501
Conc: 6.04 ng/ml