

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060624\
 Data File : PL090003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2024 21:39
 Operator : AR\AJ
 Sample : P2591-17
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BHBW3

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/06/2024
 Supervised By : Ankita Jodhani 06/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 04:24:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060424CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 04 15:06:25 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.792	11091670	10976204	5.751m	8.085 #
27)	SA Decachlor...	9.041	7.936	114.5E6	58595170	86.773m	42.261 #
Target Compounds							
9)	A Endosulfan I	6.109	5.149	246.1E6	425.8E6	120.412m	283.532m#
12)	B 4,4'-DDE	6.199	5.256	39100064	90544735	20.724	59.126 #
13)	MA Dieldrin	6.358	5.390	221.2E6	335.8E6	105.092	202.682m#
16)	A 4,4'-DDD	6.699	5.809	87375746	48055151	57.399m	38.102 #
17)	MA 4,4'-DDT	7.029	6.060	164.9E6	154.7E6	106.500	114.607

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060624\
Data File : PL090003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2024 21:39
Operator : AR\AJ
Sample : P2591-17
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 04:24:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060424CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 04 15:06:25 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

