

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062118\
 Data File : PL037279.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2018 06:20
 Operator : UA\AJ
 Sample : J3594-07
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 07:48:12 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 21 02:27:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.336	3.937	83028178	30752988	10.496	11.115
28)	SA Decachlor...	8.018	8.949	59551576	21926707	9.869	11.052
Target Compounds							
2)	A alpha-BHC	3.722f	0.000	24507262	0	2.002	N.D. #
6)	B beta-BHC	4.256f	0.000	12104261	0	2.654	N.D. #
10)	B gamma-Chl...	5.165f	0.000	15183574	0	1.621	N.D. #
23)	Chlordane-1	4.128	0.000	14376936	0	46.822	N.D. #
25)	Chlordane-3	5.165	0.000	15183574	0	14.285	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062118\
Data File : PL037279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2018 06:20
Operator : UA\AJ
Sample : J3594-07
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 07:48:12 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
Quant Title : GC Extractables
QLast Update : Thu Jun 21 02:27:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.336 min
Delta R.T.: 0.000 min
Response: 83028178
Conc: 10.50 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.937 min
Delta R.T.: -0.001 min
Response: 30752988
Conc: 11.12 ng/ml




#2 alpha-BHC

R.T.: 3.722 min
Delta R.T.: -0.046 min
Response: 24507262
Conc: 2.00 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 4.317 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.256 min
Delta R.T.: -0.043 min
Response: 12104261
Conc: 2.65 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.758 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 5.165 min
Delta R.T.: -0.068 min
Response: 15183574
Conc: 1.62 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 6.023 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 4.128 min
Delta R.T.: -0.053 min
Response: 14376936
Conc: 46.82 ng/ml

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 4.921 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.165 min
Delta R.T.: -0.067 min
Response: 15183574
Conc: 14.28 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T. : 6.022 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.018 min
Delta R.T.: -0.003 min
Response: 59551576
Conc: 9.87 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.949 min
Delta R.T.: -0.002 min
Response: 21926707
Conc: 11.05 ng/ml