

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070222\
 Data File : PL076232.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2022 16:52
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/05/2022
 Supervised By :Ankita Jodhani 07/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 22:53:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.639	4.629	42226263	105.9E6	55.831	55.845
28) SA Decachlor...	11.632	10.369	36158908	86644504	57.462	56.414
Target Compounds						
2) A alpha-BHC	6.224	5.331	56358958	141.0E6	57.057	57.680
3) MA gamma-BHC...	6.621	5.751	52744384	130.1E6	56.974	56.276
4) MA Heptachlor	7.276	6.158	51004467	125.2E6	55.601	55.184
5) MB Aldrin	7.649	6.482	46972469	114.7E6	56.588	55.567
6) B beta-BHC	6.874	6.130	24220508	56152800	55.302	53.831
7) B delta-BHC	7.145	6.390	49562677	130.0E6	63.476	57.529
8) B Heptachlo...	8.117	7.055	45922710	104.8E6	57.676	54.638
9) A Endosulfan I	8.525	7.457	40639308	95261015	56.218	54.255
10) B gamma-Chl...	8.387	7.332	43361695	105.6E6	55.787	54.917
11) B alpha-Chl...	8.469	7.400	43306203	104.2E6	55.766	54.466
12) B 4,4'-DDE	8.655	7.612	37733790	94454837	55.827	55.649
13) MA Dieldrin	8.816	7.741	40999272	102.2E6	55.846	56.800m
14) MA Endrin	9.057	8.031	34414223	85329932	55.195	56.150m
15) B Endosulfa...	9.290	8.343	34862384	80463557	54.948	53.235
16) A 4,4'-DDD	9.197	8.196	32365640	78362505	57.079	56.498
17) MA 4,4'-DDT	9.517	8.456	34855182	81247266	57.096	54.934
18) B Endrin al...	9.426	8.530	28372138	65325617	54.929	54.721
19) B Endosulfa...	9.664	8.760	35780843	84421224	55.813	55.025
20) A Methoxychlor	10.002	9.049	19690472	45058159	57.405	52.911
21) B Endrin ke...	10.166	9.277	42698780	92497577	59.240	55.987m
22) Mirex	10.638	9.464	30357478	71646643	56.442	53.540

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070222\
Data File : PL076232.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2022 16:52
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/05/2022
Supervised By : Ankita Jodhani 07/05/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 22:53:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

