

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071025\
 Data File : PL096330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 21:25
 Operator : AR\AJ
 Sample : Q2489-03
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 G3(0-6)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 02:39:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 07 15:22:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.538	2.829	67622600	95941919	18.532	16.624
28)	SA Decachlor...	9.020	7.996	50937036	80015815	18.107	16.047
Target Compounds							
4)	MA Heptachlor	4.928f	0.000	10836990	0	2.366	N.D. #
27)	Chlordane-5	6.837	0.000	4482411	0	29.697	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071025\
Data File : PL096330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 21:25
Operator : AR\AJ
Sample : Q2489-03
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
G3(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 02:39:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
Quant Title : GC Extractables
QLast Update : Mon Jul 07 15:22:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.538 min
Delta R.T.: 0.000 min
Response: 67622600
Conc: 18.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
G3(0-6)




#1 Tetrachloro-m-xylene

R.T.: 2.829 min
Delta R.T.: 0.000 min
Response: 95941919
Conc: 16.62 ng/ml




#4 Heptachlor

R.T.: 4.928 min
Delta R.T.: 0.023 min
Response: 10836990
Conc: 2.37 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T. : 4.016 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 6.837 min
Delta R.T.: -0.004 min
Response: 4482411
Conc: 29.70 ng/ml

Instrument :
ECD_L
ClientSampleId :
G3(0-6)




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.013 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.020 min
Delta R.T.: 0.000 min
Response: 50937036
Conc: 18.11 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.996 min
Delta R.T.: -0.001 min
Response: 80015815
Conc: 16.05 ng/ml