

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071918\
 Data File : PL038307.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2018 19:15
 Operator : AJ\SJ
 Sample : J3828-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 04:39:36 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.327	3.929	136.8E6	57931145	18.376	22.070
28)	SA Decachlor...	7.997	8.936	126.6E6	51037113	16.682	20.879 #
Target Compounds							
2)	A alpha-BHC	0.000	4.317	0	821591	N.D.	0.212 #
3)	MA gamma-BHC...	4.083f	4.549f	75126532	1070527	7.166	0.296 #
6)	B beta-BHC	4.243f	0.000	10787098	0	2.285	N.D. #
7)	B delta-BHC	0.000	4.960	0	3483841	N.D.	1.050 #
8)	B Heptachlo...	0.000	5.801	0	144143	N.D.	0.051 #
15)	B Endosulfa...	6.032f	6.827	30360948	890670	3.829	0.371 #
16)	A 4,4'-DDD	0.000	6.676f	0	8079028	N.D.	3.774 #
17)	MA 4,4'-DDT	0.000	7.003	0	780917	N.D.	0.366 #
23)	Chlordane-1	0.000	4.960	0	3483841	N.D.	30.063 #
27)	Chlordane-5	0.000	6.827	0	890670	N.D.	9.943 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071918\
Data File : PL038307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2018 19:15
Operator : AJ\SJ
Sample : J3828-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 04:39:36 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.327 min
Delta R.T.: -0.002 min
Response: 136835133
Conc: 18.38 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.929 min
Delta R.T.: 0.000 min
Response: 57931145
Conc: 22.07 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.758 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.317 min
Delta R.T.: 0.008 min
Response: 821591
Conc: 0.21 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.083 min
Delta R.T.: 0.049 min
Response: 75126532
Conc: 7.17 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.549 min
Delta R.T.: -0.039 min
Response: 1070527
Conc: 0.30 ng/ml




#6 beta-BHC

R.T.: 4.243 min
Delta R.T.: -0.047 min
Response: 10787098
Conc: 2.29 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 4.750 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 4.484 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.960 min
Delta R.T.: -0.005 min
Response: 3483841
Conc: 1.05 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.997 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.801 min
Delta R.T.: 0.019 min
Response: 144143
Conc: 0.05 ng/ml



#15 Endosulfan II


R.T.: 6.032 min
Delta R.T.: -0.064 min
Response: 30360948
Conc: 3.83 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.827 min
Delta R.T.: 0.022 min
Response: 890670
Conc: 0.37 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 5.956 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.676 min
Delta R.T.: -0.048 min
Response: 8079028
Conc: 3.77 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.190 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.003 min
Delta R.T.: -0.020 min
Response: 780917
Conc: 0.37 ng/ml




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T.: 4.169 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 4.960 min
Delta R.T.: 0.047 min
Response: 3483841
Conc: 30.06 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 5.883 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 6.827 min
Delta R.T.: -0.055 min
Response: 890670
Conc: 9.94 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.997 min
Delta R.T.: -0.006 min
Response: 126550376
Conc: 16.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.936 min
Delta R.T.: -0.001 min
Response: 51037113
Conc: 20.88 ng/ml