

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082119\
 Data File : PL051690.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2019 11:21
 Operator : SG\AJ
 Sample : K4383-03MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-1359MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 03:49:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.315	3.926	179.1E6	46870501	20.931	21.453
28)	SA Decachlor...	7.975	8.953	230.9E6	59486883	19.026	19.543
Target Compounds							
2)	A alpha-BHC	3.742	4.314	799.7E6	170.7E6	55.191	52.968
3)	MA gamma-BHC...	4.018	4.597	763.9E6	169.0E6	58.062	53.669
4)	MA Heptachlor	4.301	5.098	863.9E6	189.4E6	57.063	54.010
5)	MB Aldrin	4.539	5.400	839.5E6	185.1E6	58.096	56.446
6)	B beta-BHC	4.272	4.774	327.2E6	82703337	57.297	56.746
7)	B delta-BHC	4.465	4.987	665.7E6	136.4E6	44.768	43.580
8)	B Heptachlo...	4.975	5.786	793.8E6	177.0E6	56.760	52.445
9)	A Endosulfan I	5.305	6.142	771.8E6	169.1E6	57.412	52.311
10)	B gamma-Chl...	5.197	6.020	838.8E6	186.3E6	56.175	51.023
11)	B alpha-Chl...	5.253	6.092	814.1E6	182.4E6	56.253	52.856
12)	B 4,4'-DDE	5.420	6.253	844.1E6	190.5E6	61.675	54.111
13)	MA Dieldrin	5.544	6.398	852.2E6	180.7E6	56.047	52.776
14)	MA Endrin	5.796	6.615	795.5E6	163.1E6	54.540	53.918
15)	B Endosulfa...	6.071	6.827	727.6E6	160.5E6	52.289	48.698
16)	A 4,4'-DDD	5.931	6.743	683.8E6	146.6E6	55.839	54.136
17)	MA 4,4'-DDT	6.165	7.040	700.4E6	155.3E6	54.930	53.032
18)	B Endrin al...	6.238	6.952	593.4E6	138.8E6	51.011	49.118
19)	B Endosulfa...	6.450	7.175	677.0E6	157.1E6	52.040	50.859
20)	A Methoxychlor	6.709	7.501	359.9E6	85515769	53.083	50.987
21)	B Endrin ke...	6.931	7.643	744.6E6	163.4E6	52.633	51.071
22)	Mirex	7.104	8.091	527.3E6	125.0E6	49.808	51.381

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082119\
Data File : PL051690.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2019 11:21
Operator : SG\AJ
Sample : K4383-03MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-1359MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 03:49:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

