

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082721\
 Data File : PL069511.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Aug 2021 08:36
 Operator : AR\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 23:42:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 21 02:23:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

4) MA	Heptachlor	6.154f	7.197	865961	3584482	1.250	1.967 #
6) B	beta-BHC	6.154	6.792	865961	12296900	N.D.	13.085
8) B	Heptachlo...	0.000	8.036	0	7004741	N.D.	4.657 #
9) A	Endosulfan I	0.000	8.435	0	307089	N.D.	0.222 #
14) MA	Endrin	8.080f	0.000	1173472	0	2.032	N.D. #
23)	Chlordane-1	0.000	6.934f	0	3305284	N.D.	58.386 #
25)	Chlordane-3	0.000	8.330f	0	8156993	N.D.	38.091 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082721\
Data File : PL069511.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Aug 2021 08:36
Operator : AR\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 23:42:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
Quant Title : GC Extractables
QLast Update : Sat Aug 21 02:23:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 4.634 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 5.567 min
Response: 0
Conc: N.D.




#4 Heptachlor

R.T.: 6.154 min
Delta R.T.: -0.020 min
Response: 865961
Conc: 1.25 ng/ml



#4 Heptachlor


R.T.: 7.197 min
Delta R.T.: 0.004 min
Response: 3584482
Conc: 1.97 ng/ml



#6 beta-BHC


R.T.: 6.154 min
Delta R.T.: 0.011 min
Response: 865961
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#6 beta-BHC

R.T.: 6.792 min
Delta R.T.: -0.004 min
Response: 12296900
Conc: 13.08 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 7.075 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 8.036 min
Delta R.T.: 0.006 min
Response: 7004741
Conc: 4.66 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 7.478 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#9 Endosulfan I

R.T.: 8.435 min
Delta R.T.: -0.001 min
Response: 307089
Conc: 0.22 ng/ml




#14 Endrin

R.T.: 8.080 min
Delta R.T.: 0.026 min
Response: 1173472
Conc: 2.03 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 8.964 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 5.964 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#23 Chlordane-1

R.T.: 6.934 min
Delta R.T.: -0.021 min
Response: 3305284
Conc: 58.39 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 7.354 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 8.330 min
Delta R.T.: 0.027 min
Response: 8156993
Conc: 38.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.401 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 11.488 min
Response: 0
Conc: N.D.