

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082721\
 Data File : PL069525.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Aug 2021 15:12
 Operator : AR\AJ
 Sample : M3463-33RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SWCS-27-0002

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 23:46:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 21 02:23:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.633	5.568	7641636	24846981	16.563	17.345
28)	SA Decachlor...	10.401	11.492	8270923	17232726	12.254	14.061
Target Compounds							
7)	B delta-BHC	0.000	7.040f	0	2326053	N.D.	1.259 #
11)	B alpha-Chl...	0.000	8.385	0	736437	N.D.	0.493 #
16)	A 4,4'-DDD	0.000	9.110	0	960179	N.D.	0.783 #
17)	MA 4,4'-DDT	8.476	9.396f	652900	708193	1.134	0.578 #
26)	Chlordane-4	0.000	8.385	0	736437	N.D.	2.861 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082721\
Data File : PL069525.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Aug 2021 15:12
Operator : AR\AJ
Sample : M3463-33RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SWCS-27-0002

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 23:46:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
Quant Title : GC Extractables
QLast Update : Sat Aug 21 02:23:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.633 min
Delta R.T.: -0.001 min
Response: 7641636
Conc: 16.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
SWCS-27-0002




#1 Tetrachloro-m-xylene

R.T.: 5.568 min
Delta R.T.: 0.001 min
Response: 24846981
Conc: 17.35 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 6.404 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 7.040 min
Delta R.T.: -0.024 min
Response: 2326053
Conc: 1.26 ng/ml



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 7.422 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
SWCS-27-0002




#11 alpha-Chlordane

R.T.: 8.385 min
Delta R.T.: 0.004 min
Response: 736437
Conc: 0.49 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 8.220 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 9.110 min
Delta R.T.: 0.004 min
Response: 960179
Conc: 0.78 ng/ml



#17 4,4'-DDT


R.T.: 8.476 min
Delta R.T.: -0.005 min
Response: 652900
Conc: 1.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
SWCS-27-0002




#17 4,4'-DDT

R.T.: 9.396 min
Delta R.T.: -0.028 min
Response: 708193
Conc: 0.58 ng/ml




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 7.423 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 8.385 min
Delta R.T.: 0.000 min
Response: 736437
Conc: 2.86 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.401 min
Delta R.T.: 0.000 min
Response: 8270923
Conc: 12.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
SWCS-27-0002



#28 Decachlorobiphenyl

R.T.: 11.492 min
Delta R.T.: 0.004 min
Response: 17232726
Conc: 14.06 ng/ml