

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
 Data File : PL052307.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2019 18:08
 Operator : SG\AJ
 Sample : PB122924BL
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB122924BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 03:45:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.922	204.3E6	52090405	23.869	23.842
28)	SA Decachlor...	7.968	8.946	220.4E6	51473833	18.158	16.911
Target Compounds							
6)	B beta-BHC	0.000	4.799f	0	531052	N.D.	0.364 #
8)	B Heptachlo...	0.000	5.799	0	1098754	N.D.	0.326 #
11)	B alpha-Chl...	0.000	6.094	0	2295927	N.D.	0.665 #
13)	MA Dieldrin	0.000	6.403	0	4582493	N.D.	1.338 #
26)	Chlordane-4	0.000	6.094	0	2295927	N.D.	4.227 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
Data File : PL052307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2019 18:08
Operator : SG\AJ
Sample : PB122924BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB122924BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 03:45:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.312 min
Delta R.T.: -0.004 min
Response: 204289312
Conc: 23.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122924BL




#1 Tetrachloro-m-xylene

R.T.: 3.922 min
Delta R.T.: -0.004 min
Response: 52090405
Conc: 23.84 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.273 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.799 min
Delta R.T.: 0.023 min
Response: 531052
Conc: 0.36 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 4.975 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PB122924BL




#8 Heptachlor epoxide

R.T.: 5.799 min
Delta R.T.: 0.012 min
Response: 1098754
Conc: 0.33 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 5.253 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 6.094 min
Delta R.T.: 0.001 min
Response: 2295927
Conc: 0.67 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T. : 5.544 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PB122924BL




#13 Dieldrin

R.T.: 6.403 min
Delta R.T.: 0.004 min
Response: 4582493
Conc: 1.34 ng/ml




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 5.258 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 6.094 min
Delta R.T.: -0.002 min
Response: 2295927
Conc: 4.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.968 min
Delta R.T.: -0.007 min
Response: 220401158
Conc: 18.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122924BL



#28 Decachlorobiphenyl

R.T.: 8.946 min
Delta R.T.: -0.009 min
Response: 51473833
Conc: 16.91 ng/ml