

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
 Data File : PL052304.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2019 17:26
 Operator : SG\AJ
 Sample : K4736-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-GF-02-045-B

Manual Integrations
 APPROVED

Ankita
 9/10/2019 11:15:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 03:44:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.922	108.2E6	28106647	12.642	12.865
28)	SA Decachlor...	7.968	8.945	98388060	22967368	8.106	7.545
Target Compounds							
4)	MA Heptachlor	4.297	5.094	75821855	16139619	5.008	4.602
10)	B gamma-Chl...	5.193	6.016	152.3E6	59793503	10.200	16.377 #
11)	B alpha-Chl...	5.247	6.089	262.0E6	107.2E6	18.104	31.075 #
12)	B 4,4'-DDE	5.415	6.246	60642115	18078812	4.431	5.136
16)	A 4,4'-DDD	5.923	6.739	13614205	9040476	1.112	3.339m#
17)	MA 4,4'-DDT	6.158	7.032	93714360	18297781	7.349	6.250m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
Data File : PL052304.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2019 17:26
Operator : SG\AJ
Sample : K4736-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
FK-GF-02-045-B

Manual Integrations
APPROVED

Ankita
9/10/2019 11:15:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 03:44:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

