

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
 Data File : PL052329.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2019 23:18
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 9/10/2019 11:18:33 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 03:54:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.922	462.4E6	112.7E6	54.031	51.582
28)	SA Decachlor...	7.967	8.945	572.4E6	128.9E6	47.158	42.343
Target Compounds							
2)	A alpha-BHC	3.739	4.310	785.8E6	166.9E6	54.231	51.801
3)	MA gamma-BHC...	4.014	4.593	751.4E6	162.7E6	57.115	51.667
4)	MA Heptachlor	4.297	5.094	829.5E6	177.5E6	54.796	50.606
5)	MB Aldrin	4.535	5.395	823.3E6	172.3E6	56.973	52.535
6)	B beta-BHC	4.269	4.771	313.9E6	77379893	54.969	53.094
7)	B delta-BHC	4.462	4.983	826.6E6	166.4E6	55.594	53.158
8)	B Heptachlor...	4.970	5.782	766.5E6	174.3E6	54.810	51.649
9)	A Endosulfan I	5.301	6.136	729.7E6	168.3E6	54.276	52.064
10)	B gamma-Chl...	5.192	6.015	824.9E6	181.1E6	55.243	49.610
11)	B alpha-Chl...	5.248	6.086	801.3E6	181.2E6	55.371	52.499
12)	B 4,4'-DDE	5.415	6.247	684.8E6	183.1E6	50.039	52.015
13)	MA Dieldrin	5.539	6.392	839.5E6	180.3E6	55.215	52.651
14)	MA Endrin	5.791	6.609	772.1E6	154.0E6	52.932	50.924
15)	B Endosulfa...	6.067	6.822	755.1E6	163.0E6	54.268	49.465
16)	A 4,4'-DDD	5.925	6.738	659.4E6	144.0E6	53.847	53.207
17)	MA 4,4'-DDT	6.159	7.033	678.2E6	147.4E6	53.190	50.336m
18)	B Endrin al...	6.233	6.946	622.4E6	139.2E6	53.504	49.264
19)	B Endosulfa...	6.445	7.169	680.7E6	153.7E6	52.329	49.776
20)	A Methoxychlor	6.703	7.495	345.4E6	79049351	50.946	47.132
21)	B Endrin ke...	6.925	7.637	746.4E6	161.0E6	52.763	50.333
22)	Mirex	7.098	8.084	536.5E6	118.3E6	50.677	48.617

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090919\
Data File : PL052329.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2019 23:18
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
9/10/2019 11:18:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 03:54:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

