

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101122\
 Data File : PL078277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2022 12:56
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/12/2022
 Supervised By : Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 00:40:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100322.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 03 19:46:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.413	2.621	91381716	145.2E6	49.547	51.646
2)	SA Decachlor...	8.900	7.715	80805844	112.9E6	51.472	52.013
Target Compounds							
2)	A alpha-BHC	3.878	3.113	137.9E6	230.0E6	49.681	52.697
3)	MA gamma-BHC...	4.213	3.436	126.3E6	198.9E6	48.641	51.929
4)	MA Heptachlor	4.789	3.767	122.7E6	197.5E6	46.836	52.324
5)	MB Aldrin	5.129	4.041	113.1E6	188.3E6	47.711	52.768
6)	B beta-BHC	4.427	3.737	53475946	81240278	46.639	50.825
7)	B delta-BHC	4.671	3.959	104.9E6	203.8E6	44.340	52.097
8)	B Heptachlo...	5.562	4.540	94359257	163.5E6	43.938	51.502
9)	A Endosulfan I	5.949	4.905	93436967	151.9E6	48.597	52.056
10)	B gamma-Chl...	5.818	4.789	100.3E6	163.2E6	48.184	51.175
11)	B alpha-Chl...	5.895	4.852	99229314	158.2E6	48.093	51.658
12)	B 4,4'-DDE	6.074	5.045	94092018	152.5E6	48.653	50.324
13)	MA Dieldrin	6.226	5.168	95199078	161.9E6	49.284	52.483
14)	MA Endrin	6.458	5.441	79552447	123.2E6	46.729	47.391
15)	B Endosulfa...	6.686	5.737	75862488	125.4E6	41.836	47.666
16)	A 4,4'-DDD	6.597	5.596	70965483	117.8E6	48.659m	49.605m
17)	MA 4,4'-DDT	6.909	5.846	77126072	115.5E6	46.631	47.321
18)	B Endrin al...	6.817	5.916	61653487	97808472	48.194	49.757
19)	B Endosulfa...	7.051	6.138	77045309	128.3E6	49.610	51.553
20)	A Methoxychlor	7.390	6.423	39285536	64276740	47.591	49.824
21)	B Endrin ke...	7.538	6.638	82480625	139.6E6	49.321	50.822
22)	Mirex	7.998	6.813	62470142	102.1E6	49.020	52.024

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL101122\
 Data File : PL078277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2022 12:56
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/12/2022

Supervised By :Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 00:40:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL100322.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 03 19:46:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

