

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112522\
 Data File : PL079436.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2022 14:52
 Operator : AR\AJ
 Sample : N5766-04MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CR-4MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 23:57:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 24 00:09:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	2.565	41655270	54018958	18.439	19.845
2)	SA Decachlor...	8.769	7.631	36415894	46195387	18.751	19.664
Target Compounds							
2)	A alpha-BHC	3.778	3.051	174.7E6	222.9E6	50.798	51.701
3)	MA gamma-BHC...	4.111	3.372	159.5E6	201.7E6	49.213	51.791
4)	MA Heptachlor	4.682	3.698	156.7E6	188.8E6	47.569	49.069
5)	MB Aldrin	5.020	3.969	160.5E6	206.2E6	51.684	55.547
6)	B beta-BHC	4.327	3.672	66403725	74868865	47.030	49.413
7)	B delta-BHC	4.567	3.891	141.9E6	202.0E6	51.134	52.516
8)	B Heptachlo...	5.454	4.466	143.6E6	175.3E6	50.282	51.943
9)	A Endosulfan I	5.838	4.829	131.8E6	165.0E6	49.555	51.569
10)	B gamma-Chl...	5.709	4.714	145.9E6	177.8E6	50.982	52.539
11)	B alpha-Chl...	5.786	4.776	143.4E6	173.3E6	49.778	51.751
12)	B 4,4'-DDE	5.969	4.970	126.6E6	166.3E6	50.680	54.227
13)	MA Dieldrin	6.115	5.090	140.9E6	182.6E6	51.381	54.180
14)	MA Endrin	6.346	5.361	123.7E6	164.4E6	51.136	54.344
15)	B Endosulfa...	6.576	5.658	117.8E6	154.1E6	52.168	55.086
16)	A 4,4'-DDD	6.495	5.522	104.5E6	132.9E6	52.673	53.113
17)	MA 4,4'-DDT	6.804	5.768	100.9E6	126.7E6	44.497	48.485
18)	B Endrin al...	6.708	5.836	98845313	121.1E6	54.576	56.911
19)	B Endosulfa...	6.941	6.059	117.5E6	146.6E6	50.671	51.962
20)	A Methoxychlor	7.288	6.345	54319217	65293217	45.026	48.039
21)	B Endrin ke...	7.427	6.555	124.7E6	162.8E6	50.541	52.845
22)	Mirex	7.884	6.727	97247807	122.8E6	51.342	51.919

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112522\
Data File : PL079436.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2022 14:52
Operator : AR\AJ
Sample : N5766-04MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
CR-4MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 23:57:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112322.M
Quant Title : GC Extractables
QLast Update : Thu Nov 24 00:09:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

