

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112720\
 Data File : PL063255.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2020 09:01
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 15:02:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 24 01:53:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.410	3.950	88071785	164.6E6	51.713	50.444
2)	SA Decachlor...	8.126	8.989	89671770	114.6E6	48.018	48.164
Target Compounds							
2)	A alpha-BHC	3.842	4.337	129.0E6	245.4E6	52.241	50.425
3)	MA gamma-BHC...	4.121	4.619	115.6E6	227.9E6	50.584	49.848
4)	MA Heptachlor	4.418	5.124	105.8E6	206.4E6	50.388	49.529
5)	MB Aldrin	4.662	5.427	97053747	196.7E6	51.845	49.826
6)	B beta-BHC	4.371	4.791	48174740	93546705	51.537	45.880
7)	B delta-BHC	4.572	5.006	108.0E6	204.6E6	50.598	53.726
8)	B Heptachlo...	5.101	5.813	93367783	179.6E6	51.591	51.563
9)	A Endosulfan I	5.436	6.168	86962946	164.3E6	51.372	48.651
10)	B gamma-Chl...	5.326	6.047	92592749	176.2E6	51.137	49.807
11)	B alpha-Chl...	5.384	6.119	92905436	175.4E6	51.376	48.595
12)	B 4,4'-DDE	5.549	6.278	85280593	172.0E6	51.461	49.017
13)	MA Dieldrin	5.677	6.425	90332181	167.9E6	50.959	48.857
14)	MA Endrin	5.932	6.642	75551995	131.2E6	44.989	44.436
15)	B Endosulfa...	6.204	6.852	83675898	140.8E6	45.526	42.596
16)	A 4,4'-DDD	6.061	6.768	76723229	136.1E6	53.027	50.934
17)	MA 4,4'-DDT	6.300	7.066	63192637	111.9E6	46.986	43.835
18)	B Endrin al...	6.372	6.976	72612285	121.7E6	48.731	49.150
19)	B Endosulfa...	6.584	7.199	82283528	130.4E6	49.990	47.908
20)	A Methoxychlor	6.844	7.526	39844117	59304606	46.503	43.249
21)	B Endrin ke...	7.069	7.668	99872116	139.5E6	53.375	50.284
22)	Mirex	7.256	8.121	74117771	90954283	49.843	46.563

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112720\
Data File : PL063255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2020 09:01
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 15:02:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
Quant Title : GC Extractables
QLast Update : Tue Nov 24 01:53:58 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

