

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112720\
 Data File : PL063256.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2020 09:27
 Operator : DD\AJ
 Sample : L4857-01RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 304EM-STOCKPILERE

Manual Integrations
 APPROVED

Ankita
 11/30/2020 11:18:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 15:16:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 24 01:53:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.416	3.952	13488503	24296491	7.920	7.444
28)	SA Decachlor...	8.134	8.995	21371127	26351126	11.444	11.075
Target Compounds							
10)	B gamma-Chl...	5.335	6.049	18750300	50372300	10.355	14.237m#
11)	B alpha-Chl...	5.391	6.123	37473255	98017436	20.723m	27.163m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL112720\
Data File : PL063256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2020 09:27
Operator : DD\AJ
Sample : L4857-01RE
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
304EM-STOCKPILERE


Manual Integrations
APPROVED

Ankita
11/30/2020 11:18:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 15:16:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL112320.M
Quant Title : GC Extractables
QLast Update : Tue Nov 24 01:53:58 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.416 min
Delta R.T.: 0.006 min
Response: 13488503
Conc: 7.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
304EM-STOCKPILERE


Manual Integrations
APPROVED

Ankita
11/30/2020 11:18:42 AM




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.002 min
Response: 24296491
Conc: 7.44 ng/ml




#10 gamma-Chlordane

R.T.: 5.335 min
Delta R.T.: 0.008 min
Response: 18750300
Conc: 10.36 ng/ml



#10 gamma-Chlordane

R.T.: 6.049 min
Delta R.T.: 0.004 min
Response: 50372300
Conc: 14.24 ng/ml m




#11 alpha-Chlordane

R.T.: 5.391 min
Delta R.T.: 0.006 min
Response: 37473255
Conc: 20.72 ng/ml m

Instrument :
ECD_L
ClientSampleId :
304EM-STOCKPILERE


Manual Integrations
APPROVED

Ankita
11/30/2020 11:18:42 AM




#11 alpha-Chlordane

R.T.: 6.123 min
Delta R.T.: 0.004 min
Response: 98017436
Conc: 27.16 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.134 min
Delta R.T.: 0.008 min
Response: 21371127
Conc: 11.44 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.995 min
Delta R.T.: 0.005 min
Response: 26351126
Conc: 11.08 ng/ml