

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0010220\
 Data File : P0065246.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jan 2020 20:31
 Operator : AJ/MA
 Sample : K6484-50
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 30544-46-66

Manual Integrations
 APPROVED

mohammad
 1/3/2020 10:16:21 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 22:55:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0123119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 31 07:42:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.166	3.440	632332	734454	29.980	29.072
2) SA Decachlor...	9.676	8.352	737144	874987	24.155m	22.736
Target Compounds						
26) L6 AR-1254-1	6.141	5.291	647774	447742	497.860m	191.501 #
27) L6 AR-1254-2	6.378	5.437	532689	466973	251.426	219.057
28) L6 AR-1254-3	6.741	5.835	710247	721999	296.483	200.386 #
29) L6 AR-1254-4	7.024	6.063	499828	501032	268.473	215.453
30) L6 AR-1254-5	7.440	6.476	585432	892031	288.545	264.472

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010220\
Data File : PO065246.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jan 2020 20:31
Operator : AJ/MA
Sample : K6484-50
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
30544-46-66

Manual Integrations
APPROVED

mohammad
1/3/2020 10:16:21 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 22:55:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO123119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 31 07:42:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

