

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010220\
 Data File : P0065236.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jan 2020 17:44
 Operator : AJ/MA
 Sample : PB125840BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB125840BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 18:04:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0123119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 31 07:42:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.166	3.440	461105	533247	21.862	21.108
2) SA Decachlor...	9.675	8.351	704608	856626	23.089	22.259

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010220\
Data File : P0065236.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jan 2020 17:44
Operator : AJ/MA
Sample : PB125840BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB125840BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 18:04:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0123119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 31 07:42:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.166 min
Delta R.T.: -0.013 min
Response: 461105
Conc: 21.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125840BL




#1 Tetrachloro-m-xylene

R.T.: 3.440 min
Delta R.T.: -0.012 min
Response: 533247
Conc: 21.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.675 min
Delta R.T.: -0.024 min
Response: 704608
Conc: 23.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.351 min
Delta R.T.: -0.016 min
Response: 856626
Conc: 22.26 ng/ml