

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010225\
 Data File : P0108896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2025 08:57
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO010225

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 10:47:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 03 04:48:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.700	466.1E6	273.3E6	50.323	49.294
2) SA Decachlor...	8.768	8.720	378.3E6	176.8E6	53.279	47.111
Target Compounds						
21) L5 AR-1248-1	4.799	4.785	102.7E6	56203490	488.830	472.934
22) L5 AR-1248-2	5.038	5.022	141.2E6	79213981	489.044	476.314
23) L5 AR-1248-3	5.254	5.065	172.6E6	84111087	485.209	472.940
24) L5 AR-1248-4	5.608	5.236	242.3E6	97266333	484.218	470.816
25) L5 AR-1248-5	5.650	5.629	169.4E6	93999587	487.129	474.367

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010225\
Data File : PO108896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2025 08:57
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO010225

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 10:47:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 03 04:48:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

