

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010225\
 Data File : P0108879.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2025 19:54
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 01:51:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 03 01:48:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.704	3.701	698.9E6	420.1E6	72.323	72.739
2) SA Decachlor...	8.768	8.719	547.7E6	281.4E6	71.897	71.889
Target Compounds						
21) L5 AR-1248-1	4.799	4.787	149.2E6	84823176	700.456	708.098
22) L5 AR-1248-2	5.038	5.024	203.6E6	117.1E6	698.614	701.566
23) L5 AR-1248-3	5.254	5.066	250.7E6	124.6E6	699.272	701.659
24) L5 AR-1248-4	5.609	5.237	356.3E6	145.7E6	708.036	703.664
25) L5 AR-1248-5	5.651	5.630	248.6E6	141.5E6	708.363	714.607



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010225\
Data File : PO108879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2025 19:54
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 01:51:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 03 01:48:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.704 min
Delta R.T.: 0.000 min
Response: 698885182
Conc: 72.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.701 min
Delta R.T.: 0.000 min
Response: 420082650
Conc: 72.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.768 min
Delta R.T.: 0.000 min
Response: 547689624
Conc: 71.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.719 min
Delta R.T.: -0.001 min
Response: 281367810
Conc: 71.89 ng/ml



#21 AR-1248-1


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 149236457
Conc: 700.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.787 min
Delta R.T.: 0.000 min
Response: 84823176
Conc: 708.10 ng/ml




#22 AR-1248-2

R.T.: 5.038 min
Delta R.T.: 0.000 min
Response: 203552219
Conc: 698.61 ng/ml



#22 AR-1248-2


R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 117116088
Conc: 701.57 ng/ml



#23 AR-1248-3


R.T.: 5.254 min
Delta R.T.: 0.001 min
Response: 250685201
Conc: 699.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.066 min
Delta R.T.: 0.000 min
Response: 124601915
Conc: 701.66 ng/ml




#24 AR-1248-4

R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 356294381
Conc: 708.04 ng/ml



#24 AR-1248-4


R.T.: 5.237 min
Delta R.T.: 0.000 min
Response: 145691585
Conc: 703.66 ng/ml



#25 AR-1248-5

R.T.: 5.651 min
Delta R.T.: 0.000 min
Response: 248634449
Conc: 708.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.630 min
Delta R.T.: 0.000 min
Response: 141504656
Conc: 714.61 ng/ml