

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010225\
 Data File : P0108888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2025 22:39
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/03/2025
 Supervised By :Ankita Jodhani 01/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 02:24:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 03 02:23:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.700	462.2E6	274.9E6	50.000	50.000
2) SA Decachlor...	8.768	8.719	363.3E6	183.4E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	6.854	6.824	328.2E6	181.5E6	500.000	500.000
37) L8 AR-1262-2	7.355	7.323	562.5E6	304.6E6	500.000	500.000
38) L8 AR-1262-3	7.641	7.608	229.9E6	116.5E6	498.348m	500.000
39) L8 AR-1262-4	7.705	7.671	401.8E6	214.8E6	500.000	500.000
40) L8 AR-1262-5	8.203	8.165	181.7E6	92163361	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010225\
Data File : PO108888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2025 22:39
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2025
Supervised By :Ankita Jodhani 01/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 02:24:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 03 02:23:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: 0.000 min
Response: 462175000
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2025
Supervised By :Ankita Jodhani 01/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.700 min
Delta R.T.: 0.000 min
Response: 274896720
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.768 min
Delta R.T.: 0.000 min
Response: 363258502
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.719 min
Delta R.T.: 0.000 min
Response: 183373209
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 6.854 min
Delta R.T.: -0.011 min
Response: 328186528
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2025
Supervised By :Ankita Jodhani 01/03/2025




#36 AR-1262-1

R.T.: 6.824 min
Delta R.T.: -0.015 min
Response: 181460498
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 7.355 min
Delta R.T.: -0.012 min
Response: 562498337
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.323 min
Delta R.T.: -0.015 min
Response: 304556174
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 7.641 min
Delta R.T.: -0.012 min
Response: 229900288
Conc: 498.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2025
Supervised By :Ankita Jodhani 01/03/2025




#38 AR-1262-3

R.T.: 7.608 min
Delta R.T.: -0.016 min
Response: 116493476
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 7.705 min
Delta R.T.: -0.012 min
Response: 401817507
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 7.671 min
Delta R.T.: -0.016 min
Response: 214835756
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 8.203 min
Delta R.T.: -0.013 min
Response: 181661662
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/03/2025
Supervised By :Ankita Jodhani 01/03/2025



#40 AR-1262-5

R.T.: 8.165 min
Delta R.T.: -0.016 min
Response: 92163361
Conc: 500.00 ng/ml