

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010324\
 Data File : P0100894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2024 23:53
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 04 03:17:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.643	151.8E6	98789736	48.641	48.345
2) SA Decachlor...	10.091	8.702	101.1E6	78466061	48.400	51.986
Target Compounds						
16) L4 AR-1242-1	5.516	4.737	35046412	25197975	496.897	482.459
17) L4 AR-1242-2	5.538	4.757	50620448	34715787	482.272	479.010
18) L4 AR-1242-3	5.599	4.935	32857997	18761814	505.472	489.654
19) L4 AR-1242-4	5.699	5.019	25765082	20770632	514.507	498.317
20) L4 AR-1242-5	6.438	5.547	25359803	23496953	489.980	511.976



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010324\
Data File : PO100894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2024 23:53
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 04 03:17:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.347 min
Delta R.T.: -0.002 min
Response: 151771458
Conc: 48.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: -0.003 min
Response: 98789736
Conc: 48.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.091 min
Delta R.T.: -0.001 min
Response: 101113390
Conc: 48.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: -0.004 min
Response: 78466061
Conc: 51.99 ng/ml



#16 AR-1242-1


R.T.: 5.516 min
Delta R.T.: -0.003 min
Response: 35046412
Conc: 496.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.737 min
Delta R.T.: -0.006 min
Response: 25197975
Conc: 482.46 ng/ml




#17 AR-1242-2

R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 50620448
Conc: 482.27 ng/ml



#17 AR-1242-2


R.T.: 4.757 min
Delta R.T.: -0.005 min
Response: 34715787
Conc: 479.01 ng/ml



#18 AR-1242-3


R.T.: 5.599 min
Delta R.T.: -0.004 min
Response: 32857997
Conc: 505.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.935 min
Delta R.T.: -0.006 min
Response: 18761814
Conc: 489.65 ng/ml




#19 AR-1242-4

R.T.: 5.699 min
Delta R.T.: -0.002 min
Response: 25765082
Conc: 514.51 ng/ml



#19 AR-1242-4


R.T.: 5.019 min
Delta R.T.: -0.006 min
Response: 20770632
Conc: 498.32 ng/ml



#20 AR-1242-5

R.T.: 6.438 min
Delta R.T.: -0.002 min
Response: 25359803
Conc: 489.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.547 min
Delta R.T.: -0.006 min
Response: 23496953
Conc: 511.98 ng/ml