

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010325\
 Data File : P0108902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jan 2025 10:48
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 12:25:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 03 04:48:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.704	3.700	461.3E6	270.6E6	49.803	48.808
2)	SA Decachlor...	8.769	8.719	375.5E6	173.4E6	52.886	46.197
Target Compounds							
26)	L6 AR-1254-1	5.609	5.589	253.4E6	141.6E6	479.403	467.216
27)	L6 AR-1254-2	5.758	5.736	220.7E6	125.8E6	473.391	464.009
28)	L6 AR-1254-3	6.165	6.140	355.5E6	200.7E6	472.424	468.544
29)	L6 AR-1254-4	6.394	6.368	217.0E6	119.4E6	506.663	500.262
30)	L6 AR-1254-5	6.815	6.786	311.0E6	173.5E6	481.318	468.303



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010325\
Data File : PO108902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jan 2025 10:48
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 12:25:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 03 04:48:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.704 min
Delta R.T.: 0.000 min
Response: 461325199
Conc: 49.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.700 min
Delta R.T.: 0.000 min
Response: 270610180
Conc: 48.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.769 min
Delta R.T.: 0.000 min
Response: 375462666
Conc: 52.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.719 min
Delta R.T.: 0.000 min
Response: 173371446
Conc: 46.20 ng/ml



#26 AR-1254-1


R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 253380367
Conc: 479.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 141606095
Conc: 467.22 ng/ml




#27 AR-1254-2

R.T.: 5.758 min
Delta R.T.: 0.000 min
Response: 220726743
Conc: 473.39 ng/ml



#27 AR-1254-2


R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 125780181
Conc: 464.01 ng/ml



#28 AR-1254-3


R.T.: 6.165 min
Delta R.T.: 0.000 min
Response: 355505089
Conc: 472.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 200687819
Conc: 468.54 ng/ml




#29 AR-1254-4

R.T.: 6.394 min
Delta R.T.: 0.000 min
Response: 217047384
Conc: 506.66 ng/ml



#29 AR-1254-4


R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 119430144
Conc: 500.26 ng/ml



#30 AR-1254-5

R.T.: 6.815 min
Delta R.T.: 0.000 min
Response: 310978892
Conc: 481.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.786 min
Delta R.T.: 0.000 min
Response: 173520410
Conc: 468.30 ng/ml