

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010421\
 Data File : P0074353.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jan 2021 23:06
 Operator : DD\AJ
 Sample : AR1262ICC050
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC050

Manual Integrations
 APPROVED

Ankita
 1/5/2021 1:53:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 04:21:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 05 04:19:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.932	3.989	469496	289680	4.663	5.215
2) SA Decachlor...	10.954	9.249	527753	288752	5.096	4.951
Target Compounds						
36) L8 AR-1262-1	8.583	7.341	426233	137306	50.541	53.089
37) L8 AR-1262-2	9.141	7.891	737306	480526	50.575	48.517
38) L8 AR-1262-3	9.450	8.174	467826	202349	48.310	50.370m
39) L8 AR-1262-4	9.544	8.239	377051	331080	51.873	48.906
40) L8 AR-1262-5	10.210	8.733	245446	146863	47.585	47.258

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010421\
Data File : PO074353.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Jan 2021 23:06
Operator : DD\AJ
Sample : AR1262ICC050
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleID :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
1/5/2021 1:53:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 04:21:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 04:19:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.932 min
Delta R.T.: -0.001 min
Response: 469496
Conc: 4.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
1/5/2021 1:53:03 PM




#1 Tetrachloro-m-xylene

R.T.: 3.989 min
Delta R.T.: 0.000 min
Response: 289680
Conc: 5.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.954 min
Delta R.T.: 0.000 min
Response: 527753
Conc: 5.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.249 min
Delta R.T.: 0.000 min
Response: 288752
Conc: 4.95 ng/ml




#36 AR-1262-1

R.T.: 8.583 min
Delta R.T.: 0.000 min
Response: 426233
Conc: 50.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
1/5/2021 1:53:03 PM




#36 AR-1262-1

R.T.: 7.341 min
Delta R.T.: 0.000 min
Response: 137306
Conc: 53.09 ng/ml




#37 AR-1262-2

R.T.: 9.141 min
Delta R.T.: 0.000 min
Response: 737306
Conc: 50.57 ng/ml



#37 AR-1262-2

R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 480526
Conc: 48.52 ng/ml




#38 AR-1262-3

R.T.: 9.450 min
Delta R.T.: 0.000 min
Response: 467826
Conc: 48.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050


Manual Integrations
APPROVED

Ankita
1/5/2021 1:53:03 PM




#38 AR-1262-3

R.T.: 8.174 min
Delta R.T.: -0.001 min
Response: 202349
Conc: 50.37 ng/ml m




#39 AR-1262-4

R.T.: 9.544 min
Delta R.T.: 0.000 min
Response: 377051
Conc: 51.87 ng/ml



#39 AR-1262-4

R.T.: 8.239 min
Delta R.T.: 0.000 min
Response: 331080
Conc: 48.91 ng/ml




#40 AR-1262-5

R.T.: 10.210 min
Delta R.T.: 0.000 min
Response: 245446
Conc: 47.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC050

Manual Integrations
APPROVED

Ankita
1/5/2021 1:53:03 PM



#40 AR-1262-5

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 146863
Conc: 47.26 ng/ml