

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:00:05 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091809.D 1000 =P0091807.D 750 =P0091808.D
 250 =P0091810.D 50 =P0091811.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.311	3.126	3.216	3.301	3.185	3.228	E6	2.43
2) SA	Decachlorobiphenyl	2.426	2.192	2.284	2.518	2.541	2.392	E6	6.30
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	7.050	6.347	6.742	7.370	7.229	6.948	E4	5.90
22) L5	AR-1248-2	1.049	0.936	0.995	1.105	1.067	1.031	E5	6.42
23) L5	AR-1248-3	1.133	1.035	1.077	1.200	1.266	1.142	E5	8.12
24) L5	AR-1248-4	1.227	1.118	1.166	1.274	1.253	1.208	E5	5.35
25) L5	AR-1248-5	1.200	1.095	1.141	1.240	1.224	1.180	E5	5.13
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0091809.D 1000 =P0091807.D 750 =P0091808.D
 250 =P0091810.D 50 =P0091811.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.503	1.405	1.453	1.500	1.449	1.462	E6	2.81
2) SA	Decachlorobiphenyl	1.283	1.141	1.203	1.360	1.477	1.293	E6	10.21
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:00:05 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091809.D 1000 =P0091807.D 750 =P0091808.D
 250 =P0091810.D 50 =P0091811.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	3.286	2.772	2.933	3.478	3.150	3.124 E4	8.95
22) L5	AR-1248-2	4.851	4.233	4.510	5.203	5.194	4.798 E4	8.88
23) L5	AR-1248-3	5.095	4.449	4.737	5.448	5.235	4.993 E4	8.00
24) L5	AR-1248-4	5.859	5.148	5.467	6.226	6.278	5.796 E4	8.41
25) L5	AR-1248-5	5.495	4.886	5.160	5.796	5.595	5.387 E4	6.73
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range