

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:17:48 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091814.D 1000 =P0091812.D 750 =P0091813.D
 250 =P0091815.D 50 =P0091816.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	3.300	3.184	3.228	3.383	2.906	3.201	E6	5.65
2)	SA	Decachlorobiphenyl	2.397	2.230	2.305	2.563	2.530	2.405	E6	5.92
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	1.253	1.150	1.193	1.353	1.446	1.279	E5	9.43
27)	L6	AR-1254-2	1.879	1.738	1.796	2.022	2.488	1.985	E5	15.16
28)	L6	AR-1254-3	1.912	1.799	1.844	2.021	2.351	1.985	E5	11.11
29)	L6	AR-1254-4	1.380	1.299	1.333	1.465	1.641	1.424	E5	9.61
30)	L6	AR-1254-5	1.482	1.392	1.434	1.570	1.484	1.473	E5	4.53
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0091814.D 1000 =P0091812.D 750 =P0091813.D
 250 =P0091815.D 50 =P0091816.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.487	1.435	1.457	1.534	1.385	1.460 E6	3.83	
2) SA	Decachlorobiphenyl	1.269	1.169	1.210	1.378	1.458	1.297 E6	9.24	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:17:48 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091814.D 1000 =P0091812.D 750 =P0091813.D
 250 =P0091815.D 50 =P0091816.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	8.399	7.733	8.009	9.127	8.953	8.444 E4	7.06
27) L6	AR-1254-2	7.419	6.756	7.026	8.087	9.215	7.701 E4	12.77
28) L6	AR-1254-3	1.155	1.069	1.105	1.239	1.422	1.198 E5	11.72
29) L6	AR-1254-4	7.012	6.461	6.676	7.561	9.503	7.442 E4	16.45
30) L6	AR-1254-5	1.025	0.943	0.979	1.069	1.280	1.059 E5	12.50
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range