

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:43:48 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091820.D 1000 =P0091818.D 750 =P0091819.D
 250 =P0091821.D 50 =P0091822.D

Compound			500	1000	750	250	50	Avg		%RSD
1)	SA	Tetrachloro-m-xylene	3.298	3.237	3.292	3.383	3.096	3.261	E6	3.26
2)	SA	Decachlorobiphenyl	4.213	3.974	4.097	4.444	4.234	4.192	E6	4.17
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1	4.054					4.054	E4	0.00
9)	L2	AR-1221-2	2.982					2.982	E4	0.00
10)	L2	AR-1221-3	8.997					8.997	E4	0.00
11)	L3	AR-1232-1	7.745					7.745	E4	0.00
12)	L3	AR-1232-2	3.808					3.808	E4	0.00
13)	L3	AR-1232-3	7.098					7.098	E4	0.00
14)	L3	AR-1232-4	3.482					3.482	E4	0.00
15)	L3	AR-1232-5	2.736					2.736	E4	0.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1	1.660					1.660	E5	0.00
37)	L8	AR-1262-2	2.777					2.777	E5	0.00
38)	L8	AR-1262-3	1.951					1.951	E5	0.00
39)	L8	AR-1262-4	9.766					9.766	E4	0.00
40)	L8	AR-1262-5	1.037					1.037	E5	0.00
41)	L9	AR-1268-1	3.645	3.493	3.572	3.794	3.644	3.630	E5	3.06
42)	L9	AR-1268-2	3.308	3.172	3.239	3.455	3.369	3.309	E5	3.34
43)	L9	AR-1268-3	2.820	2.709	2.767	2.942	2.678	2.783	E5	3.75
44)	L9	AR-1268-4	1.233	1.181	1.205	1.289	1.213	1.224	E5	3.33
45)	L9	AR-1268-5	8.854	8.661	8.803	9.030	8.363	8.742	E5	2.86

Signal #2 Calibration Files

500 =P0091820.D 1000 =P0091818.D 750 =P0091819.D
 250 =P0091821.D 50 =P0091822.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.493	1.446	1.473	1.518	1.351	1.456	E6	4.43
2) SA	Decachlorobiphenyl	2.190	2.046	2.112	2.346	2.410	2.221	E6	6.94
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0010423.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 05 01:43:48 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0091820.D 1000 =P0091818.D 750 =P0091819.D
 250 =P0091821.D 50 =P0091822.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	1.826					1.826 E4	0.00
9) L2	AR-1221-2	1.383					1.383 E4	0.00
10) L2	AR-1221-3	4.174					4.174 E4	0.00
11) L3	AR-1232-1	3.659					3.659 E4	0.00
12) L3	AR-1232-2	3.281					3.281 E4	0.00
13) L3	AR-1232-3	1.767					1.767 E4	0.00
14) L3	AR-1232-4	1.738					1.738 E4	0.00
15) L3	AR-1232-5	1.894					1.894 E4	0.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	9.863					9.863 E4	0.00
37) L8	AR-1262-2	1.633					1.633 E5	0.00
38) L8	AR-1262-3	6.876					6.876 E4	0.00
39) L8	AR-1262-4	1.214					1.214 E5	0.00
40) L8	AR-1262-5	5.785					5.785 E4	0.00
41) L9	AR-1268-1	2.104	1.987	2.041	2.220	2.253	2.121 E5	5.37
42) L9	AR-1268-2	1.872	1.773	1.817	1.967	1.951	1.876 E5	4.46
43) L9	AR-1268-3	1.640	1.547	1.589	1.732	1.762	1.654 E5	5.54
44) L9	AR-1268-4	6.859	6.314	6.556	7.332	7.252	6.863 E4	6.38
45) L9	AR-1268-5	4.671	4.518	4.595	4.813	4.681	4.655 E5	2.36

(#) = Out of Range