

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010423\
 Data File : P0091815.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2023 20:52
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 01:14:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 05 01:06:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.292	3.469	84583406	38348184	25.834	25.943
2) SA Decachlor...	10.035	8.465	64064685	34454399	26.991	27.425
Target Compounds						
26) L6 AR-1254-1	6.328	5.355	33820671	22816682	273.380	274.339
27) L6 AR-1254-2	6.549	5.504	50555311	20216332	271.962	276.106
28) L6 AR-1254-3	6.917	5.908	50518433	30966767	266.721	271.206
29) L6 AR-1254-4	7.204	6.138	36623335	18902488	267.502	272.864
30) L6 AR-1254-5	7.626	6.557	39262070	26730812	267.146	266.260



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010423\
Data File : PO091815.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2023 20:52
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 01:14:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010423.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 05 01:06:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.292 min
Delta R.T.: 0.000 min
Response: 84583406
Conc: 25.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.469 min
Delta R.T.: 0.000 min
Response: 38348184
Conc: 25.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.035 min
Delta R.T.: 0.000 min
Response: 64064685
Conc: 26.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.465 min
Delta R.T.: 0.000 min
Response: 34454399
Conc: 27.42 ng/ml



#26 AR-1254-1


R.T.: 6.328 min
Delta R.T.: 0.000 min
Response: 33820671
Conc: 273.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 22816682
Conc: 274.34 ng/ml




#27 AR-1254-2

R.T.: 6.549 min
Delta R.T.: 0.000 min
Response: 50555311
Conc: 271.96 ng/ml



#27 AR-1254-2


R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 20216332
Conc: 276.11 ng/ml



#28 AR-1254-3


R.T.: 6.917 min
Delta R.T.: 0.000 min
Response: 50518433
Conc: 266.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 5.908 min
Delta R.T.: 0.000 min
Response: 30966767
Conc: 271.21 ng/ml




#29 AR-1254-4

R.T.: 7.204 min
Delta R.T.: 0.000 min
Response: 36623335
Conc: 267.50 ng/ml



#29 AR-1254-4


R.T.: 6.138 min
Delta R.T.: 0.000 min
Response: 18902488
Conc: 272.86 ng/ml



#30 AR-1254-5

R.T.: 7.626 min
Delta R.T.: 0.000 min
Response: 39262070
Conc: 267.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.557 min
Delta R.T.: 0.000 min
Response: 26730812
Conc: 266.26 ng/ml