

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010524\
 Data File : P0100929.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2024 14:08
 Operator : YP/AJ
 Sample : P1027-03 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC24857-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 23:28:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.643	5919788	4053896	1.897	1.984
2) SA Decachlor...	10.086	8.700	3931616	2969291	1.882	1.967
Target Compounds						
3) L1 AR-1016-1	5.506	4.742	313410	418909	3.652	6.769 #
4) L1 AR-1016-2	5.564	4.755	174447	127074	1.367	1.441
5) L1 AR-1016-3	5.616	4.949	36863	92977	0.467	1.990 #
6) L1 AR-1016-4	5.703	4.986	20356	221413	0.334	5.395 #
7) L1 AR-1016-5	6.000	5.197	28057	474979	0.458	9.144 #
8) L2 AR-1221-1	4.532	3.869	148204	215587	4.009	8.585 #
9) L2 AR-1221-2	4.662	3.945	377481	186930	13.592	9.735 #
10) L2 AR-1221-3	4.731	4.057	90533	121001	1.085	2.056 #
11) L3 AR-1232-1	4.731	4.057	90533	121001	1.358	2.559 #
12) L3 AR-1232-2	5.262	4.755	97482	127074	2.781	3.047
13) L3 AR-1232-3	5.564	4.949	174447	92977	2.924	4.334 #
14) L3 AR-1232-4	5.703	5.014	20356	232343	0.742	10.935 #
15) L3 AR-1232-5	5.815	5.197	279531	474979	11.626	21.735 #
16) L4 AR-1242-1	5.506	4.742	313410	418909	4.444	8.021 #
17) L4 AR-1242-2	5.564	4.755	174447	127074	1.662	1.753
18) L4 AR-1242-3	5.616	4.949	36863	92977	0.567	2.427 #
19) L4 AR-1242-4	5.703	5.014	20356	232343	0.406	5.574 #
20) L4 AR-1242-5	6.443	5.579	192930	533925	3.728	11.634 #
21) L5 AR-1248-1	5.506	4.742	313410	418909	5.713	10.476 #
22) L5 AR-1248-2	5.815	4.986	279531	221413	3.237	3.680
23) L5 AR-1248-3	6.000	5.014	28057	232343	0.320	3.717 #
24) L5 AR-1248-4	6.417	5.197	503270	474979	5.598	6.594
25) L5 AR-1248-5	6.443	5.598	192930	463347	2.149	7.511 #
26) L6 AR-1254-1	6.417f	5.579	503270	533925	4.788	5.202
27) L6 AR-1254-2	6.576	5.694	22988	331556	0.156	3.609 #
28) L6 AR-1254-3	6.963	6.084	637302	537227	4.551	3.972
29) L6 AR-1254-4	7.212f	6.343	239167	537716	2.813	7.455 #
30) L6 AR-1254-5	7.626f	6.754	4655570	358516	39.345	3.111 #
31) L7 AR-1260-1	7.091	6.248	554726	652041	4.049	6.845 #
32) L7 AR-1260-2	7.395	6.414	29686	441959	0.236	4.112 #
33) L7 AR-1260-3	7.747	6.591	106394	150896	1.098	1.472 #
34) L7 AR-1260-4	7.950	7.042	655369	94611	6.210	1.160 #
35) L7 AR-1260-5	8.307	7.302	3260146	106512	15.444	0.590 #
36) L8 AR-1262-1	7.626f	6.754	4655570	358516	46.602	5.747 #
37) L8 AR-1262-2	8.307	7.042	3260146	94611	13.738	0.837 #
38) L8 AR-1262-3	8.600	7.652f	149294	720895	0.919	9.361 #
39) L8 AR-1262-4	8.687	7.652	119103	720895	0.924	4.880 #
40) L8 AR-1262-5	9.336	8.128	81178	248428	1.018	3.885 #
41) L9 AR-1268-1	8.592	7.652f	298830	720895	1.058	3.252 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010524\
 Data File : P0100929.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2024 14:08
 Operator : YP/AJ
 Sample : P1027-03 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC24857-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 23:28:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.687	7.652	119103	720895	0.471	3.564 #
43) L9	AR-1268-3	8.921	7.848	54056	355097	0.237	1.882 #
44) L9	AR-1268-4	9.336	8.128	81178	248428	1.037	3.700 #
45) L9	AR-1268-5	9.761	8.416	145300	899918	0.203	1.735 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010524\
Data File : PO100929.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2024 14:08
Operator : YP/AJ
Sample : P1027-03 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 23:28:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: -0.003 min
Response: 5919788
Conc: 1.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: -0.004 min
Response: 4053896
Conc: 1.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.086 min
Delta R.T.: -0.006 min
Response: 3931616
Conc: 1.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.700 min
Delta R.T.: -0.006 min
Response: 2969291
Conc: 1.97 ng/ml



#3 AR-1016-1


R.T.: 5.506 min
Delta R.T.: -0.011 min
Response: 313410
Conc: 3.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#3 AR-1016-1

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 418909
Conc: 6.77 ng/ml




#4 AR-1016-2

R.T.: 5.564 min
Delta R.T.: 0.025 min
Response: 174447
Conc: 1.37 ng/ml



#4 AR-1016-2


R.T.: 4.755 min
Delta R.T.: -0.007 min
Response: 127074
Conc: 1.44 ng/ml



#5 AR-1016-3


R.T.: 5.616 min
Delta R.T.: 0.015 min
Response: 36863
Conc: 0.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#5 AR-1016-3

R.T.: 4.949 min
Delta R.T.: 0.009 min
Response: 92977
Conc: 1.99 ng/ml




#6 AR-1016-4

R.T.: 5.703 min
Delta R.T.: 0.003 min
Response: 20356
Conc: 0.33 ng/ml



#6 AR-1016-4


R.T.: 4.986 min
Delta R.T.: 0.004 min
Response: 221413
Conc: 5.40 ng/ml



#7 AR-1016-5


R.T.: 6.000 min
Delta R.T.: 0.004 min
Response: 28057
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#7 AR-1016-5

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 474979
Conc: 9.14 ng/ml




#8 AR-1221-1

R.T.: 4.532 min
Delta R.T.: -0.026 min
Response: 148204
Conc: 4.01 ng/ml



#8 AR-1221-1


R.T.: 3.869 min
Delta R.T.: 0.007 min
Response: 215587
Conc: 8.59 ng/ml



#9 AR-1221-2


R.T.: 4.662 min
Delta R.T.: 0.018 min
Response: 377481
Conc: 13.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#9 AR-1221-2

R.T.: 3.945 min
Delta R.T.: -0.003 min
Response: 186930
Conc: 9.73 ng/ml




#10 AR-1221-3

R.T.: 4.731 min
Delta R.T.: 0.010 min
Response: 90533
Conc: 1.08 ng/ml



#10 AR-1221-3


R.T.: 4.057 min
Delta R.T.: 0.032 min
Response: 121001
Conc: 2.06 ng/ml



#11 AR-1232-1


R.T.: 4.731 min
Delta R.T.: 0.011 min
Response: 90533
Conc: 1.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#11 AR-1232-1

R.T.: 4.057 min
Delta R.T.: 0.031 min
Response: 121001
Conc: 2.56 ng/ml




#12 AR-1232-2

R.T.: 5.262 min
Delta R.T.: 0.012 min
Response: 97482
Conc: 2.78 ng/ml



#12 AR-1232-2


R.T.: 4.755 min
Delta R.T.: -0.007 min
Response: 127074
Conc: 3.05 ng/ml



#13 AR-1232-3


R.T.: 5.564 min
Delta R.T.: 0.024 min
Response: 174447
Conc: 2.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#13 AR-1232-3

R.T.: 4.949 min
Delta R.T.: 0.009 min
Response: 92977
Conc: 4.33 ng/ml




#14 AR-1232-4

R.T.: 5.703 min
Delta R.T.: 0.002 min
Response: 20356
Conc: 0.74 ng/ml



#14 AR-1232-4


R.T.: 5.014 min
Delta R.T.: -0.010 min
Response: 232343
Conc: 10.94 ng/ml



#15 AR-1232-5


R.T.: 5.815 min
Delta R.T.: 0.022 min
Response: 279531
Conc: 11.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#15 AR-1232-5

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 474979
Conc: 21.73 ng/ml




#16 AR-1242-1

R.T.: 5.506 min
Delta R.T.: -0.012 min
Response: 313410
Conc: 4.44 ng/ml



#16 AR-1242-1


R.T.: 4.742 min
Delta R.T.: -0.001 min
Response: 418909
Conc: 8.02 ng/ml



#17 AR-1242-2


R.T.: 5.564 min
Delta R.T.: 0.024 min
Response: 174447
Conc: 1.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#17 AR-1242-2

R.T.: 4.755 min
Delta R.T.: -0.008 min
Response: 127074
Conc: 1.75 ng/ml




#18 AR-1242-3

R.T.: 5.616 min
Delta R.T.: 0.013 min
Response: 36863
Conc: 0.57 ng/ml



#18 AR-1242-3


R.T.: 4.949 min
Delta R.T.: 0.008 min
Response: 92977
Conc: 2.43 ng/ml



#19 AR-1242-4


R.T.: 5.703 min
Delta R.T.: 0.002 min
Response: 20356
Conc: 0.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#19 AR-1242-4

R.T.: 5.014 min
Delta R.T.: -0.010 min
Response: 232343
Conc: 5.57 ng/ml




#20 AR-1242-5

R.T.: 6.443 min
Delta R.T.: 0.003 min
Response: 192930
Conc: 3.73 ng/ml



#20 AR-1242-5


R.T.: 5.579 min
Delta R.T.: 0.026 min
Response: 533925
Conc: 11.63 ng/ml



#21 AR-1248-1


R.T.: 5.506 min
Delta R.T.: -0.011 min
Response: 313410
Conc: 5.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#21 AR-1248-1

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 418909
Conc: 10.48 ng/ml




#22 AR-1248-2

R.T.: 5.815 min
Delta R.T.: 0.023 min
Response: 279531
Conc: 3.24 ng/ml



#22 AR-1248-2


R.T.: 4.986 min
Delta R.T.: 0.004 min
Response: 221413
Conc: 3.68 ng/ml



#23 AR-1248-3


R.T.: 6.000 min
Delta R.T.: 0.004 min
Response: 28057
Conc: 0.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#23 AR-1248-3

R.T.: 5.014 min
Delta R.T.: -0.009 min
Response: 232343
Conc: 3.72 ng/ml




#24 AR-1248-4

R.T.: 6.417 min
Delta R.T.: 0.017 min
Response: 503270
Conc: 5.60 ng/ml



#24 AR-1248-4


R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 474979
Conc: 6.59 ng/ml



#25 AR-1248-5


R.T.: 6.443 min
Delta R.T.: 0.004 min
Response: 192930
Conc: 2.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#25 AR-1248-5

R.T.: 5.598 min
Delta R.T.: 0.004 min
Response: 463347
Conc: 7.51 ng/ml




#26 AR-1254-1

R.T.: 6.417 min
Delta R.T.: 0.043 min
Response: 503270
Conc: 4.79 ng/ml



#26 AR-1254-1


R.T.: 5.579 min
Delta R.T.: 0.027 min
Response: 533925
Conc: 5.20 ng/ml



#27 AR-1254-2


R.T.: 6.576 min
Delta R.T.: -0.018 min
Response: 22988
Conc: 0.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#27 AR-1254-2

R.T.: 5.694 min
Delta R.T.: -0.007 min
Response: 331556
Conc: 3.61 ng/ml




#28 AR-1254-3

R.T.: 6.963 min
Delta R.T.: 0.002 min
Response: 637302
Conc: 4.55 ng/ml



#28 AR-1254-3


R.T.: 6.084 min
Delta R.T.: -0.025 min
Response: 537227
Conc: 3.97 ng/ml



#29 AR-1254-4


R.T.: 7.212 min
Delta R.T.: -0.035 min
Response: 239167
Conc: 2.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#29 AR-1254-4

R.T.: 6.343 min
Delta R.T.: 0.005 min
Response: 537716
Conc: 7.45 ng/ml




#30 AR-1254-5

R.T.: 7.626 min
Delta R.T.: -0.043 min
Response: 4655570
Conc: 39.35 ng/ml



#30 AR-1254-5


R.T.: 6.754 min
Delta R.T.: -0.006 min
Response: 358516
Conc: 3.11 ng/ml



#31 AR-1260-1


R.T.: 7.091 min
Delta R.T.: -0.035 min
Response: 554726
Conc: 4.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#31 AR-1260-1

R.T.: 6.248 min
Delta R.T.: 0.009 min
Response: 652041
Conc: 6.85 ng/ml




#32 AR-1260-2

R.T.: 7.395 min
Delta R.T.: 0.010 min
Response: 29686
Conc: 0.24 ng/ml



#32 AR-1260-2


R.T.: 6.414 min
Delta R.T.: -0.015 min
Response: 441959
Conc: 4.11 ng/ml



#33 AR-1260-3


R.T.: 7.747 min
Delta R.T.: 0.002 min
Response: 106394
Conc: 1.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#33 AR-1260-3

R.T.: 6.591 min
Delta R.T.: 0.008 min
Response: 150896
Conc: 1.47 ng/ml




#34 AR-1260-4

R.T.: 7.950 min
Delta R.T.: -0.021 min
Response: 655369
Conc: 6.21 ng/ml



#34 AR-1260-4


R.T.: 7.042 min
Delta R.T.: -0.016 min
Response: 94611
Conc: 1.16 ng/ml



#35 AR-1260-5


R.T.: 8.307 min
Delta R.T.: 0.019 min
Response: 3260146
Conc: 15.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#35 AR-1260-5

R.T.: 7.302 min
Delta R.T.: 0.000 min
Response: 106512
Conc: 0.59 ng/ml




#36 AR-1262-1

R.T.: 7.626 min
Delta R.T.: -0.038 min
Response: 4655570
Conc: 46.60 ng/ml



#36 AR-1262-1


R.T.: 6.754 min
Delta R.T.: -0.005 min
Response: 358516
Conc: 5.75 ng/ml



#37 AR-1262-2


R.T.: 8.307 min
Delta R.T.: 0.019 min
Response: 3260146
Conc: 13.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#37 AR-1262-2

R.T.: 7.042 min
Delta R.T.: -0.016 min
Response: 94611
Conc: 0.84 ng/ml




#38 AR-1262-3

R.T.: 8.600 min
Delta R.T.: 0.002 min
Response: 149294
Conc: 0.92 ng/ml



#38 AR-1262-3


R.T.: 7.652 min
Delta R.T.: 0.066 min
Response: 720895
Conc: 9.36 ng/ml



#39 AR-1262-4


R.T.: 8.687 min
Delta R.T.: -0.001 min
Response: 119103
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#39 AR-1262-4

R.T.: 7.652 min
Delta R.T.: 0.001 min
Response: 720895
Conc: 4.88 ng/ml




#40 AR-1262-5

R.T.: 9.336 min
Delta R.T.: -0.004 min
Response: 81178
Conc: 1.02 ng/ml



#40 AR-1262-5


R.T.: 8.128 min
Delta R.T.: -0.021 min
Response: 248428
Conc: 3.88 ng/ml



#41 AR-1268-1


R.T.: 8.592 min
Delta R.T.: -0.003 min
Response: 298830
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#41 AR-1268-1

R.T.: 7.652 min
Delta R.T.: 0.066 min
Response: 720895
Conc: 3.25 ng/ml




#42 AR-1268-2

R.T.: 8.687 min
Delta R.T.: -0.003 min
Response: 119103
Conc: 0.47 ng/ml



#42 AR-1268-2


R.T.: 7.652 min
Delta R.T.: 0.000 min
Response: 720895
Conc: 3.56 ng/ml



#43 AR-1268-3


R.T.: 8.921 min
Delta R.T.: 0.004 min
Response: 54056
Conc: 0.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1




#43 AR-1268-3

R.T.: 7.848 min
Delta R.T.: -0.010 min
Response: 355097
Conc: 1.88 ng/ml




#44 AR-1268-4

R.T.: 9.336 min
Delta R.T.: -0.005 min
Response: 81178
Conc: 1.04 ng/ml



#44 AR-1268-4


R.T.: 8.128 min
Delta R.T.: -0.021 min
Response: 248428
Conc: 3.70 ng/ml



#45 AR-1268-5

R.T.: 9.761 min
Delta R.T.: 0.008 min
Response: 145300
Conc: 0.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC24857-1-1



#45 AR-1268-5

R.T.: 8.416 min
Delta R.T.: -0.029 min
Response: 899918
Conc: 1.73 ng/ml