

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010621\
 Data File : P0074411.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jan 2021 15:59
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 1/7/2021 11:14:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 16:22:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 05 05:03:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.933	3.990	6004449	3188868	62.490	59.864
2)	SA Decachlor...	10.953	9.247	4468855	2417147	48.070	46.432
Target Compounds							
3)	L1 AR-1016-1	6.256	5.235	1605950	1210056	448.230	560.387 #
4)	L1 AR-1016-2	6.279	5.255	2304181	1701170	463.836	566.738
5)	L1 AR-1016-3	6.345	5.445	1462630	913921	494.400	570.869
6)	L1 AR-1016-4	6.452	5.497	1320743	744183	531.027m	588.353
7)	L1 AR-1016-5	6.768	5.724	1064789	903754	455.229	571.490 #
31)	L7 AR-1260-1	7.948	6.812	1883459	1458079	455.228	489.567
32)	L7 AR-1260-2	8.215	7.010	2733407	1993636	471.094	489.271
33)	L7 AR-1260-3	8.581	7.162	2247629	1709251	432.964	491.026m
34)	L7 AR-1260-4	8.812	7.642	2207677	1341684	442.147	469.251
35)	L7 AR-1260-5	9.139	7.889	5273029	3537638	465.095	460.662

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010621\
Data File : PO074411.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jan 2021 15:59
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1660CCC500

Manual Integrations
APPROVED

Ankita
1/7/2021 11:14:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 16:22:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:03:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

