

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010719\
 Data File : P0053271.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jan 2019 18:01
 Operator : SM/SJ
 Sample : K1058-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-(1-10)-COMP

Manual Integrations
 APPROVED

Sohil
 1/8/2019 11:31:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 00:41:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 02 14:46:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.174	3.211	25708104	4604728	19.648m	20.521
2)	SA Decachlor...	9.632	7.853	10751957	3221803	13.570m	12.533m
Target Compounds							
31)	L7 AR-1260-1	6.890	5.567	6256253	2548829	139.524m	166.049m
32)	L7 AR-1260-2	7.143	5.751	10702513	2113142	192.698	110.418m#
33)	L7 AR-1260-3	7.495	5.887	5228383	1952181	149.318	112.861m
34)	L7 AR-1260-4	7.718	6.338	7540086	1597488	193.938	138.096 #
35)	L7 AR-1260-5	8.025	6.578	10337548	3958247	133.944	137.402

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010719\
Data File : PO053271.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jan 2019 18:01
Operator : SM/SJ
Sample : K1058-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
TP-(1-10)-COMP

Manual Integrations
APPROVED

Sohil
1/8/2019 11:31:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 00:41:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010219.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 02 14:46:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

