

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010720\
 Data File : P0065345.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jan 2020 9:43
 Operator : AJ/MA
 Sample : L1029-07
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 X068-03-G

Manual Integrations
 APPROVED

mohammad
 1/7/2020 4:16:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 10:04:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.164	3.438	598654	770167	32.784m	35.012
2)	SA Decachlor...	9.673	8.348	633585	715671	18.699	17.257
Target Compounds							
26)	L6 AR-1254-1	6.157	5.289	224581	373880	264.733m	258.078
27)	L6 AR-1254-2	6.374	5.435	561985	680300	408.951	514.314 #
28)	L6 AR-1254-3	6.737	5.834	668239	1385906	431.151	614.996 #
29)	L6 AR-1254-4	7.021	6.060	521130	575402	408.821	381.671
30)	L6 AR-1254-5	7.436	6.473	1812804	1830450	1290.876	822.328 #
31)	L7 AR-1260-1	6.899	5.961	876273	1187120	635.334	630.556
32)	L7 AR-1260-2	7.155	6.149	1416064	1582688	807.093	651.448
33)	L7 AR-1260-3	7.510	6.300	1318686	1067634	932.988	493.245 #
34)	L7 AR-1260-4	7.733	6.767	1236228	1135801	708.004	566.138
35)	L7 AR-1260-5	8.040	7.010	1751917	2566653	522.402	518.465

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010720\
Data File : PO065345.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jan 2020 9:43
Operator : AJ/MA
Sample : L1029-07
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 10:04:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.164 min
Delta R.T.: -0.003 min
Response: 598654
Conc: 32.78 ng/ml m

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#1 Tetrachloro-m-xylene

R.T.: 3.438 min
Delta R.T.: -0.004 min
Response: 770167
Conc: 35.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.673 min
Delta R.T.: -0.005 min
Response: 633585
Conc: 18.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.348 min
Delta R.T.: -0.004 min
Response: 715671
Conc: 17.26 ng/ml




#26 AR-1254-1

R.T.: 6.157 min
Delta R.T.: -0.005 min
Response: 224581
Conc: 264.73 ng/ml m

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#26 AR-1254-1

R.T.: 5.289 min
Delta R.T.: -0.004 min
Response: 373880
Conc: 258.08 ng/ml




#27 AR-1254-2

R.T.: 6.374 min
Delta R.T.: -0.004 min
Response: 561985
Conc: 408.95 ng/ml



#27 AR-1254-2

R.T.: 5.435 min
Delta R.T.: -0.004 min
Response: 680300
Conc: 514.31 ng/ml




#28 AR-1254-3

R.T.: 6.737 min
Delta R.T.: -0.004 min
Response: 668239
Conc: 431.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#28 AR-1254-3

R.T.: 5.834 min
Delta R.T.: -0.003 min
Response: 1385906
Conc: 615.00 ng/ml




#29 AR-1254-4

R.T.: 7.021 min
Delta R.T.: -0.005 min
Response: 521130
Conc: 408.82 ng/ml



#29 AR-1254-4

R.T.: 6.060 min
Delta R.T.: -0.004 min
Response: 575402
Conc: 381.67 ng/ml




#30 AR-1254-5

R.T.: 7.436 min
Delta R.T.: -0.005 min
Response: 1812804
Conc: 1290.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#30 AR-1254-5

R.T.: 6.473 min
Delta R.T.: -0.005 min
Response: 1830450
Conc: 822.33 ng/ml




#31 AR-1260-1

R.T.: 6.899 min
Delta R.T.: -0.003 min
Response: 876273
Conc: 635.33 ng/ml



#31 AR-1260-1

R.T.: 5.961 min
Delta R.T.: -0.004 min
Response: 1187120
Conc: 630.56 ng/ml




#32 AR-1260-2

R.T.: 7.155 min
Delta R.T.: -0.005 min
Response: 1416064
Conc: 807.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#32 AR-1260-2

R.T.: 6.149 min
Delta R.T.: -0.004 min
Response: 1582688
Conc: 651.45 ng/ml




#33 AR-1260-3

R.T.: 7.510 min
Delta R.T.: -0.004 min
Response: 1318686
Conc: 932.99 ng/ml



#33 AR-1260-3

R.T.: 6.300 min
Delta R.T.: -0.004 min
Response: 1067634
Conc: 493.25 ng/ml




#34 AR-1260-4

R.T.: 7.733 min
Delta R.T.: -0.005 min
Response: 1236228
Conc: 708.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
X068-03-G


Manual Integrations
APPROVED

mohammad
1/7/2020 4:16:15 PM




#34 AR-1260-4

R.T.: 6.767 min
Delta R.T.: -0.004 min
Response: 1135801
Conc: 566.14 ng/ml



#35 AR-1260-5

R.T.: 8.040 min
Delta R.T.: -0.004 min
Response: 1751917
Conc: 522.40 ng/ml



#35 AR-1260-5

R.T.: 7.010 min
Delta R.T.: -0.004 min
Response: 2566653
Conc: 518.47 ng/ml