

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010725\
 Data File : P0108942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2025 08:19
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 11:59:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 03 04:48:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.702	3.698	478.3E6	282.2E6	51.638	50.895
2) SA Decachlor...	8.769	8.719	328.4E6	155.2E6	46.250	41.357
Target Compounds						
3) L1 AR-1016-1	4.798	4.784	161.2E6	90552104	501.012	493.237
4) L1 AR-1016-2	4.817	4.804	222.9E6	126.4E6	503.637	501.462
5) L1 AR-1016-3	4.874	4.979	158.3E6	71100341	503.137	501.187
6) L1 AR-1016-4	4.994	5.021	124.0E6	58789131	501.496	495.188
7) L1 AR-1016-5	5.252	5.234	131.4E6	74962594	499.791	488.920
31) L7 AR-1260-1	6.295	6.269	234.3E6	134.4E6	499.910	498.674
32) L7 AR-1260-2	6.484	6.456	276.2E6	162.9E6	484.169	502.232
33) L7 AR-1260-3	6.853	6.610	218.4E6	149.8E6	454.850	479.677
34) L7 AR-1260-4	7.114	7.082	191.2E6	117.1E6	437.191	469.940
35) L7 AR-1260-5	7.355	7.323	454.7E6	257.5E6	451.522	455.012



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010725\
Data File : PO108942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jan 2025 08:19
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 11:59:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 03 04:48:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.702 min
Delta R.T.: -0.002 min
Response: 478326045
Conc: 51.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.698 min
Delta R.T.: -0.003 min
Response: 282186119
Conc: 50.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.769 min
Delta R.T.: 0.000 min
Response: 328353013
Conc: 46.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.719 min
Delta R.T.: 0.000 min
Response: 155204530
Conc: 41.36 ng/ml



#3 AR-1016-1


R.T.: 4.798 min
Delta R.T.: -0.002 min
Response: 161174871
Conc: 501.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.784 min
Delta R.T.: -0.002 min
Response: 90552104
Conc: 493.24 ng/ml




#4 AR-1016-2

R.T.: 4.817 min
Delta R.T.: -0.002 min
Response: 222942862
Conc: 503.64 ng/ml



#4 AR-1016-2


R.T.: 4.804 min
Delta R.T.: -0.003 min
Response: 126382291
Conc: 501.46 ng/ml



#5 AR-1016-3


R.T.: 4.874 min
Delta R.T.: -0.002 min
Response: 158315193
Conc: 503.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.979 min
Delta R.T.: -0.002 min
Response: 71100341
Conc: 501.19 ng/ml




#6 AR-1016-4

R.T.: 4.994 min
Delta R.T.: -0.002 min
Response: 123985689
Conc: 501.50 ng/ml



#6 AR-1016-4


R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 58789131
Conc: 495.19 ng/ml



#7 AR-1016-5


R.T.: 5.252 min
Delta R.T.: -0.002 min
Response: 131366124
Conc: 499.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.234 min
Delta R.T.: -0.003 min
Response: 74962594
Conc: 488.92 ng/ml




#31 AR-1260-1

R.T.: 6.295 min
Delta R.T.: -0.002 min
Response: 234270757
Conc: 499.91 ng/ml



#31 AR-1260-1


R.T.: 6.269 min
Delta R.T.: -0.002 min
Response: 134447598
Conc: 498.67 ng/ml



#32 AR-1260-2


R.T.: 6.484 min
Delta R.T.: -0.001 min
Response: 276238959
Conc: 484.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.456 min
Delta R.T.: -0.002 min
Response: 162890892
Conc: 502.23 ng/ml




#33 AR-1260-3

R.T.: 6.853 min
Delta R.T.: 0.000 min
Response: 218383160
Conc: 454.85 ng/ml



#33 AR-1260-3


R.T.: 6.610 min
Delta R.T.: -0.002 min
Response: 149827989
Conc: 479.68 ng/ml



#34 AR-1260-4


R.T.: 7.114 min
Delta R.T.: 0.000 min
Response: 191171965
Conc: 437.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.082 min
Delta R.T.: -0.002 min
Response: 117124242
Conc: 469.94 ng/ml



#35 AR-1260-5

R.T.: 7.355 min
Delta R.T.: 0.000 min
Response: 454686213
Conc: 451.52 ng/ml



#35 AR-1260-5

R.T.: 7.323 min
Delta R.T.: 0.000 min
Response: 257456262
Conc: 455.01 ng/ml