

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010821\
 Data File : P0074460.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jan 2021 17:52
 Operator : DD\AJ
 Sample : M1040-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 41006

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 00:38:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 05 05:03:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.933	3.991	1427859	837485	14.860	15.722
2) SA Decachlor...	10.952	9.246	1347522	665926	14.495	12.792

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010821\
Data File : P0074460.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2021 17:52
Operator : DD\AJ
Sample : M1040-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
41006

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 00:38:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010421.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 05 05:03:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.933 min
Delta R.T.: 0.000 min
Response: 1427859
Conc: 14.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
41006




#1 Tetrachloro-m-xylene

R.T.: 3.991 min
Delta R.T.: 0.002 min
Response: 837485
Conc: 15.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.952 min
Delta R.T.: -0.004 min
Response: 1347522
Conc: 14.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.246 min
Delta R.T.: -0.003 min
Response: 665926
Conc: 12.79 ng/ml