

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010824\
 Data File : P0100980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jan 2024 16:50
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 00:30:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.640	156.0E6	100.7E6	50.010	49.277
2) SA Decachlor...	10.082	8.695	103.1E6	77075786	49.345	51.065
Target Compounds						
21) L5 AR-1248-1	5.512	4.734	27941959	20233133	509.330	505.972
22) L5 AR-1248-2	5.786	4.972	44039878	29785189	509.913	495.078
23) L5 AR-1248-3	5.990	5.015	43241775	31158613	492.652	498.464
24) L5 AR-1248-4	6.395	5.188	44471761	35867600	494.642	497.949
25) L5 AR-1248-5	6.433	5.585	43507716	31702531	484.709	513.884



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010824\
Data File : PO100980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jan 2024 16:50
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 00:30:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.005 min
Response: 156045045
Conc: 50.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: -0.006 min
Response: 100692873
Conc: 49.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.082 min
Delta R.T.: -0.010 min
Response: 103088068
Conc: 49.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.695 min
Delta R.T.: -0.011 min
Response: 77075786
Conc: 51.07 ng/ml



#21 AR-1248-1


R.T.: 5.512 min
Delta R.T.: -0.006 min
Response: 27941959
Conc: 509.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.734 min
Delta R.T.: -0.009 min
Response: 20233133
Conc: 505.97 ng/ml




#22 AR-1248-2

R.T.: 5.786 min
Delta R.T.: -0.006 min
Response: 44039878
Conc: 509.91 ng/ml



#22 AR-1248-2


R.T.: 4.972 min
Delta R.T.: -0.009 min
Response: 29785189
Conc: 495.08 ng/ml



#23 AR-1248-3


R.T.: 5.990 min
Delta R.T.: -0.006 min
Response: 43241775
Conc: 492.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.015 min
Delta R.T.: -0.008 min
Response: 31158613
Conc: 498.46 ng/ml




#24 AR-1248-4

R.T.: 6.395 min
Delta R.T.: -0.006 min
Response: 44471761
Conc: 494.64 ng/ml



#24 AR-1248-4


R.T.: 5.188 min
Delta R.T.: -0.009 min
Response: 35867600
Conc: 497.95 ng/ml



#25 AR-1248-5

R.T.: 6.433 min
Delta R.T.: -0.006 min
Response: 43507716
Conc: 484.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.585 min
Delta R.T.: -0.009 min
Response: 31702531
Conc: 513.88 ng/ml