

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0010924\
 Data File : P0100994.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2024 10:25
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1242CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/10/2024
 Supervised By :Ankita Jodhani 01/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 01:07:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.343	3.638	176.6E6	112.2E6	56.611	54.905
2) SA Decachlor...	10.082	8.694	109.5E6	84040214	52.399	55.680
Target Compounds						
16) L4 AR-1242-1	5.511	4.733	40421448	27108274	573.105	519.035m
17) L4 AR-1242-2	5.533	4.752	58708347	39517279	559.327	545.262
18) L4 AR-1242-3	5.595	4.931	38706838	20524630	595.448	535.661
19) L4 AR-1242-4	5.694	5.014	30248599	23170253	604.039	555.887
20) L4 AR-1242-5	6.432	5.543	30917801	26361052	597.367	574.382

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010924\
Data File : PO100994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2024 10:25
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/10/2024
Supervised By :Ankita Jodhani 01/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 01:07:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

