

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011019\
 Data File : P0053340.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jan 2019 9:50
 Operator : SM/SJ
 Sample : K1082-07
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 00:41:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 09 03:34:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.168	3.204	35534160	3886261	26.593	15.542 #
2)	SA Decachlor...	9.628	7.847	10591936	3076045	13.410	13.432
Target Compounds							
9)	L2 AR-1221-2	4.466	0.000	151502	0	13.805	N.D. #
10)	L2 AR-1221-3	0.000	3.599	0	189869	N.D.	24.824 #
11)	L3 AR-1232-1	0.000	3.599	0	189869	N.D.	31.098 #
20)	L4 AR-1242-5	0.000	4.927	0	117460	N.D.	13.267 #
25)	L5 AR-1248-5	0.000	4.927	0	117460	N.D.	10.597 #
26)	L6 AR-1254-1	0.000	4.927	0	117460	N.D.	5.952 #
27)	L6 AR-1254-2	6.362	0.000	399396	0	6.239	N.D. #
28)	L6 AR-1254-3	6.733	0.000	301699	0	4.601	N.D. #
29)	L6 AR-1254-4	0.000	5.601	0	112507	N.D.	6.856 #
31)	L7 AR-1260-1	0.000	5.562	0	544451	N.D.	36.647 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011019\
Data File : PO053340.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jan 2019 9:50
Operator : SM/SJ
Sample : K1082-07
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 00:41:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010819.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 09 03:34:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.168 min
Delta R.T.: -0.003 min
Response: 35534160
Conc: 26.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.204 min
Delta R.T.: -0.003 min
Response: 3886261
Conc: 15.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.628 min
Delta R.T.: -0.006 min
Response: 10591936
Conc: 13.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.847 min
Delta R.T.: -0.004 min
Response: 3076045
Conc: 13.43 ng/ml



#9 AR-1221-2


R.T.: 4.466 min
Delta R.T.: -0.002 min
Response: 151502
Conc: 13.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.481 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.542 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 3.599 min
Delta R.T.: 0.050 min
Response: 189869
Conc: 24.82 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.541 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.599 min
Delta R.T.: 0.050 min
Response: 189869
Conc: 31.10 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.220 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 4.927 min
Delta R.T.: -0.006 min
Response: 117460
Conc: 13.27 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 6.220 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 4.927 min
Delta R.T.: -0.041 min
Response: 117460
Conc: 10.60 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.153 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 117460
Conc: 5.95 ng/ml



#27 AR-1254-2


R.T.: 6.362 min
Delta R.T.: -0.005 min
Response: 399396
Conc: 6.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.070 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.733 min
Delta R.T.: 0.002 min
Response: 301699
Conc: 4.60 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.445 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 7.013 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 5.601 min
Delta R.T.: -0.064 min
Response: 112507
Conc: 6.86 ng/ml



#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.888 min
Response: 0
Conc: N.D.



#31 AR-1260-1

R.T.: 5.562 min
Delta R.T.: -0.004 min
Response: 544451
Conc: 36.65 ng/ml